



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**CMi INTOXILYZER 8000 MAINTENANCE REPORT**

RECEIVED  
By: [unclear] Date: 11/03/2022

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER <b>80-006195</b>	LOCATION OF INSTRUMENT <b>CARTHAGE POLICE DEPT</b>	DATE OF INSPECTION <b>11/03/2022</b>	TIME OF INSPECTION <b>03:53</b>
--	---	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	03:55	DRY	AG102002	01/20/2023
Cal Check	0.100	03:56	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	03:56	N/A	N/A	N/A
Cal Check	0.100	03:57	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	03:57	0.100	INTOXIMETERS	
Cal Check	0.099	03:57	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	03:58	0.100		
<b>Pass</b>			CALIBRATION CHECK RESULT 2		
			0.100		
			CALIBRATION CHECK RESULT 3		
			0.099		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.0%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	03:59
RAM Test	Pass		Subject Test	RFI*	03:59
EEPROM Checksum Test	Pass		Air Blank	0.000	03:59
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
<b>Pass</b>			<b>Pass</b>		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	3	0	4	4	4	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
OCT 2022 MAINT

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME <b>RANSOME</b>	
TYPE II PERMIT NUMBER <b>220138</b>	EXPIRATION DATE <b>05/11/2024</b>	TELEPHONE NUMBER <b>6207043962</b>



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 21-Jan-2021

Lot # AG102002 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
20-Jan-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2021.01.27 14:59:44 -06:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: Rod Marsala  
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JACCOB RANSOME**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

*Laura P. Nacy*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220138

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/11/2024

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator RANSOME, JACCOB  
 Permit No 220138  
 Date Issued 5/11/2022 Date Expires 5/11/2024

