

RECEIVED

By Tracy Crews at 10:45 am, Aug 19, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

Epel

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007514	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 08/08/2022	TIME OF INSPECTION 00:33
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 33321080A3	STANDARD EXPIRATION DATE 01/05/2024
Air Blank	0.000	00:35	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.078	00:35	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI INC	
Air Blank	0.000	00:36	CALIBRATION CHECK RESULT 1 0.078		
Cal Check	0.077	00:36	CALIBRATION CHECK RESULT 2 0.077		
Air Blank	0.000	00:36	CALIBRATION CHECK RESULT 3 0.078		
Cal Check	0.078	00:37	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 3.7%		
Air Blank	0.000	00:37	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS	
Voltage/Current Test	Pass	Test	g/210L
RAM Test	Pass	Air Blank	RFI*
EEPROM Checksum Test	Pass	Air Blank	0.000
Real Time Clock Test	Pass	*RFI Detect	
DSP Test	Pass		
Analytical Stability Test	Pass		
Modem Test	Pass		
Temperature Regulation Test	Pass		

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	2	4	3	6

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE <i>Po. J. Ch...</i>	PRINT NAME JORDAN INFRANCA
----------------------------------	-------------------------------

TYPE II PERMIT NUMBER 210128	EXPIRATION DATE 06/21/2023	TELEPHONE NUMBER 816-382-5897
---------------------------------	-------------------------------	----------------------------------



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 14496

Part #: BAC105L080T

Cylinder Size: 105L

Lot Number: 33321080A3

Expiration: 1/5/2024

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy		Distributed by:
		(U, n=2):	Analytical Method:	
Ethanol	288 ppm Balance	+/- 0.002 BAC(6/20L)	NDIR	CMI Inc. 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com
Nitrogen				

*Traceable to: Certified Reference Material - 262.4 µmol/mol Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11
Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]

01-04-2022
Issuance Date



The calibration results shown on this certificate were obtained using equipment and materials capable of producing certified reference materials as NIST and apply only to the stated purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of the material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

Jordan A. Nag
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DATE 6/21/2021

NUMBER 210128

EXPIRES 6/21/2023

MO 586-0771 (R-10)

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named candidate is authorized to operate an individual breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: JORDAN INFRANCA, JORDAN
Date Issued: 6/21/2021 Date Expires: 6/21/2023