

RECEIVED

By Tracy Crews at 8:22 am, May 22, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>Sparta 107985</u>	PRINTER SN <u>099.3586.820</u>	DATE OF INSPECTION <u>5/18/2023</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>200 North Ave Sparta MO 65753</u>		TIME OF INSPECTION <u>1130</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LAB LOT # 21380 EXP. DATE 9/13/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP3584 SIMULATOR EXP DATE 3/27/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .102 TEST 2 → .102 TEST 3 → .101

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

INSPECTING OFFICER

SIGNATURE Tremous Taylor

PRINT NAME Tremous Taylor

TYPE II PERMIT NUMBER/EXPIRATION DATE 230054 3/22/2025

TELEPHONE NUMBER 417 634 3992

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



SPARTA POLICE DEPARTMENT

200 North Avenue / P.O. Box 246 – Sparta, MO 65753
(417) 242 5511 (417) 634 5785 FAX



AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01372

Temp	Date	Time	s/	210L
Air Blank:				
	05/18/23	11:38	.000	
Calibration Check:				
21	05/18/23	11:38	.102	

Subject Name

Test #1
Subject I.D.
D. Taylor Col

Operator Name, I.D.

Sparta P.D.
Location

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01373

Temp	Date	Time	s/	210L
Air Blank:				
	05/18/23	11:41	.000	
Calibration Check:				
22	05/18/23	11:41	.102	

Subject Name

Test #2
Subject I.D.

Operator Name, I.D.

D. Taylor Col
Location
Sparta P.D.

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01374

Temp	Date	Time	s/	210L
Air Blank:				
	05/18/23	11:43	.000	
Calibration Check:				
22	05/18/23	11:43	.101	

Subject Name

Test #3
Subject I.D.

Operator Name, I.D.

D. Taylor Col
Location
Sparta P.D.

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01375

Temp	Date	Time	s/	210L
VOID: RFI				
	12	05/18/23	11:45	

Subject Name

RFI
Subject I.D.

Operator Name, I.D.

D. Taylor Col
Location

Sparta P.D.

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01376

Temp	Date	Time	s/	210L
Air Blank:				
	05/18/23	11:47	.000	
Subject Test: Auto				
23	05/18/23	11:47	.000	

Subject Name

Blank
Subject I.D.

Operator Name, I.D.

D. Taylor Col
Location
Sparta P.D.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP3584 Manufacturer: Guth
Model Number: 12V500
Agency: SPARTA PD
Agency Address: 200 NORTH AVE, SPARTA, MO 65753

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/24/2022 Date of Expiration: 10/24/2023

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.01), Combined Uncertainty (.03)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 3/27/2023
Certification Expiration: 3/27/2024
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none
Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: MP3584_3272023

X Brianna Medrano (signature)

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TRAMPUS TAYLOR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/27/2023

NUMBER 230054

EXPIRES 3/27/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (16-10)

MO 680-0771 (6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TAYLOR, TRAMPUS
Permit No 230054
Date Issued 3/27/2023 Date Expires 3/27/2025

