

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>030792</b>	NAME OF AGENCY <b>Missouri State Highway Patrol</b>	DATE OF INSPECTION <b>05/17/2024</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>5268 Flat River RD, Park Hills, MO 63601 (Troop C Service Center)</b>		TIME OF INSPECTION <b>10:09 pm</b>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <b>Guth Laboratories, Inc.</b>	LOT # <b>23180</b> EXP. DATE <b>05/17/2025</b>

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **33.99** SIM. SN **MP2460** SIM. NIST EXP DATE **01/16/2025**

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <b>.101</b>	TEST 2 <b>.100</b>	TEST 3 <b>.100</b>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<b>2</b> (0-.04)	(.05-.09)	<b>1</b> (.10-.14)	(.15-.19)	(OVER .19)
----------	------------------	-----------	--------------------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME <b>Zachary L. Ricker</b>
IDENTIFICATION NUMBER/EXPIRATION DATE <b>230304 12/11/2025</b>	TELEPHONE NUMBER <b>(636) 300-3800</b>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 030792  
Version no: 532B

TEST RECORD 00393  
g/  
s/

Temp Date Time 210L  
Air Blank: 05/17/24 22:09 .000  
Calibration Check: 22 05/17/24 22:09 .101

Subject Name  
Maint 1  
Subject I.D.

Operator Name, I.D.  
Z.L. Ricker 230304  
Location

5268 Flat River Rd  
Park Hills, Mo

AS IV Serial no: 030792  
Version no: 532B

TEST RECORD 00394  
g/  
s/

Temp Date Time 210L  
Air Blank: 05/17/24 22:12 .000  
Calibration Check: 23 05/17/24 22:12 .100

Subject Name  
Maint 2  
Subject I.D.

Z.L. Ricker 230304  
Operator Name, I.D.

Location  
5268 Flat River Rd

Park Hills, Mo

AS IV Serial no: 030792  
Version no: 532B

TEST RECORD 00395  
g/  
s/

Temp Date Time 210L  
Air Blank: 05/17/24 22:14 .000  
Calibration Check: 24 05/17/24 22:14 .100

Subject Name  
Maint 3  
Subject I.D.

Operator Name, I.D.  
Z.L. Ricker 230304  
Location

5268 Flat River Rd

Park Hills, Mo

AS IV Serial no: 030792  
Version no: 532B

TEST RECORD 00396  
g/  
s/

Temp Date Time 210L  
VOID: RFI  
12 05/17/24 22:15

Subject Name  
Maint RFI  
Subject I.D.

Operator Name, I.D.  
Z.L. Ricker 230304  
Location

5268 Flat River Rd

Park Hills, Mo



**GUTH LABORATORIES, INC.**

890 NORTH 6TH STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-654-6478

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23180 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1220% (w/vol) ethyl alcohol. The expiration date for this lot number is May 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**ZACHARY RICKER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2023

NUMBER 230304

EXPIRES 12/11/2025

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** RICKER, ZACHARY  
**Permit No** 230304  
**Date Issued** 12/11/2023 **Date Expires** 12/11/2025

