



**RECEIVED**

By Tracy Crews at 10:21 am, Aug 21, 2024

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030810	NAME OF AGENCY El Dorado Springs Police Dept	DATE OF INSPECTION 08/14/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 1207 South Main Street El Dorado Springs		TIME OF INSPECTION 10:22 AM

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories

**RECEIVED**  
LOT # 23380 EXP. DATE 10/17/2025  
SERIAL # MP24946 SIM. INST. EXP. DATE 08/23/2025  
By Tracy Crews at 10:21 am, Aug 21, 2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • 100

TEST 2 • 101

TEST 3 • 100

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

N/A

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Jarrod Schiereck
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>220233-09/12/2024</u>	TELEPHONE NUMBER (417) 876-2313

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 030810  
Version no: 532B

TEST RECORD 00669

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/14/24 16:22 .000  
Subject Test: Auto  
23 08/14/24 16:22 .000

Subject Name  
BLANK TEST  
Subject I.D.

Operator Name, I.D. 109  
Jared Schiereck  
Location  
1207 S MAIN  
EL DORADO SPRINGS

AS IV Serial no: 030810  
Version no: 532B

TEST RECORD 00670

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/14/24 16:23 .000  
Calibration Check:  
23 08/14/24 16:23 .100

Subject Name  
TEST #1  
Subject I.D.

Operator Name, I.D. 109  
Jared Schiereck  
Location  
1207 S MAIN  
EL DORADO SPRINGS

AS IV Serial no: 030810  
Version no: 532B

TEST RECORD 00671

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/14/24 16:26 .000  
Calibration Check:  
24 08/14/24 16:26 .101

Subject Name  
TEST #2  
Subject I.D.

Operator Name, I.D. 109  
Jared Schiereck  
Location  
1207 S MAIN  
EL DORADO SPRINGS

AS IV Serial no: 030810  
Version no: 532B

TEST RECORD 00672

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/14/24 16:27 .000  
Calibration Check:  
24 08/14/24 16:27 .100

Subject Name  
TEST #3  
Subject I.D.

Operator Name, I.D. 109  
Jared Schiereck  
Location  
1207 S MAIN  
EL DORADO SPRINGS

AS IV Serial no: 030810  
Version no: 532B

TEST RECORD 00673

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 08/14/24 16:29

Subject Name  
RFI TEST  
Subject I.D.

Operator Name, I.D. 109  
Jared Schiereck  
Location  
1207 S MAIN  
EL DORADO SPRINGS



## GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L**  $\pm 3\%$ .

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JARROD SCHIERECK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/12/2022

NUMBER 220233

EXPIRES 9/12/2024

*Mike Maxson*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator SCHIERECK, JARROD  
 Permit No 220233  
 Date Issued 9/12/2022 Date Expires 9/12/2024

