



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 7:52 am, Aug 01, 2024

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111744	NAME OF AGENCY BOONE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 08/01/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E. COUNTY DR, COLUMBIA		TIME OF INSPECTION 3:25 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG404403 EXP. DATE 02/13/2026

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .098

TEST 2 ← .096

TEST 3 ← .096

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	3	(.10-.14)	3	(.15-.19)	3	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

CLOCK ADJUSTED. INSTRUMENT OPERATING SATISFACTORILY AND WITHIN ESTABLISHED LIMITS.

INSPECTING OFFICER

SIGNATURE
Johnathan Wells

PRINT NAME
JOHNATHAN WELLS

TYPE OF PERMIT NUMBER/EXPIRATION DATE
240088, 04/05/2026

TELEPHONE NUMBER
(573) 875-1111

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111744
Version no: 532B

TEST RECORD 008880
Temp Date Time 210L
g/

Air Blank:
08/01/24 03:31 .000
Calibration Check:
20 08/01/24 03:31 .098

Subject Name

Test one

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

J. Wells 240088

Location

2111 E. County Dr

Columbia, MO

AS IV Serial no: 111744
Version no: 532B

TEST RECORD 008881
Temp Date Time 210L
g/

Air Blank:
08/01/24 03:33 .000
Calibration Check:
22 08/01/24 03:33 .096

Subject Name

Test Two

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

J. Wells 240088

Location

2111 E. County Dr

Columbia, MO

AS IV Serial no: 111744
Version no: 532B

TEST RECORD 008882
Temp Date Time 210L
g/

Air Blank:
08/01/24 03:36 .000
Calibration Check:
23 08/01/24 03:36 .096

Subject Name

Test Three

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

J. Wells 240088

Location

2111 E. County Dr

Columbia, MO

AS IV Serial no: 111744
Version no: 532B

TEST RECORD 008883
Temp Date Time 210L
g/

VOID: RFI
12 08/01/24 03:39

Subject Name

VOID RFI

Subject I.D.

Monthly Maintenance

Operator Name, I.D.

J. Wells 240088

Location

2111 E. County Dr

Columbia, MO



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 14-Feb-2024

Lot # AG404403 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
13-Feb-2026	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)


Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: **NDIR**

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:02.16.2024 09:23



Approved for Release: _____
Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JOHNATHAN WELLS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/2/2024

NUMBER 240088

EXPIRES 4/2/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula F. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WELLS, JOHNATHAN
Permit No 240088
Date Issued 4/2/2024 **Date Expires** 4/2/2026

