

CARROLLTON POLICE DEPARTMENT-BREATH ALCOHOL PROGRAM

98 IV SERIAL NO1 087971
 VERSION NO1 0805

TEST RECORD 08088

9/ /
 TEST DATE TIME 2122
 26/26/24 19147 0222
 CARROLLTON, MO
 26 09/26/24 19147 0222

SUBJECT NAME
TEST #1
 SUBJECT ID#

98 IV SERIAL NO1 087971
 VERSION NO1 0805

TEST RECORD 08088

9/ /
 TEST DATE TIME 2122
 26/26/24 19149 0226
 CARROLLTON, MO
 26 09/26/24 19149 0226

SUBJECT NAME
TEST #2
 SUBJECT ID#

98 IV SERIAL NO1 087971
 VERSION NO1 0805

TEST RECORD 08088

9/ /
 TEST DATE TIME 2122
 26/26/24 19151 0202
 CARROLLTON, MO
 26 09/26/24 19151 0202

SUBJECT NAME
TEST #3
 SUBJECT ID#

98 IV SERIAL NO1 087971
 VERSION NO1 0805

TEST RECORD 08088

9/ /
 TEST DATE TIME 2122
 12 26/26/24 19153
 CARROLLTON, MO

SUBJECT NAME
RFI TEST
 SUBJECT ID#

OPERATOR NAME: W. BARGER #230029
 ADDRESS: 1320 N US 65
 CITY: CARROLLTON, MO
 STATE: MO
 ZIP: 64633

OPERATOR NAME: W. BARGER #230029
 ADDRESS: 1320 N US 65
 CITY: CARROLLTON, MO
 STATE: MO
 ZIP: 64633

OPERATOR NAME: W. BARGER #230029
 ADDRESS: 1320 N US 65
 CITY: CARROLLTON, MO
 STATE: MO
 ZIP: 64633

OPERATOR NAME: W. BARGER #230029
 ADDRESS: 1320 N US 65
 CITY: CARROLLTON, MO
 STATE: MO
 ZIP: 64633

TEST #1

TEST #2

TEST #3

RFI TEST

AS-IV S/N: 087971 Type II Permit Holder: William Barger #230029-Exp: 02/17/2025 Date of Maintenance Report: 6/4/2024

Carrollton Police Department AS-IV Monthly Maintenance Report



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-584-8170

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 22430 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 1, 2022, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is November 30, 2024 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Certified Reference Standard lot number KN03052002 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

WILLIAM BARGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/17/2023

Mike Masoun

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230029

David T. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 2/17/2025

LAB-4 (18-10)

MO 680-0771 (8-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **BARGER, WILLIAM**
Permit No **230029**
Date Issued **2/17/2023** Date Expires **2/17/2025**

