



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 10:42 am, May 20, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087978	NAME OF AGENCY Owensville PD	DATE OF INSPECTION 04/18/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 780 West Highway 28 Owensville		TIME OF INSPECTION 2:24 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>23390</u> EXP. DATE <u>10/17/2025</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> SIM. SN <u>SD 2300</u> SIM. NIST EXP DATE <u>10/06/2024</u>	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 .103	TEST 2 .103	TEST 3 .102
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operating satisfactorily and within established limits.

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Jonathan Scott Griffith
TYPE II PERMIT NUMBER/EXPIRATION DATE 230060/04-02-2025	TELEPHONE NUMBER (573) 437-2195

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 087978
Version no: 532B

TEST RECORD 00782

Temp Date Time ^{s/} 210L

Air Blank:
04/18/24 14:24 .000
Subject Test: Auto
21 04/18/24 14:24 .000

Subject Name

Subject I.D.

Blank Test

Operator Name, I.D.

GRIFFITH 101

Location

Owensville PD

AS IV Serial no: 087978
Version no: 532B

TEST RECORD 00783

Temp Date Time ^{s/} 210L

Air Blank:
04/18/24 14:27 .000
Calibration Check:
21 04/18/24 14:27 .103

Subject Name

Subject I.D.

Check 1

Operator Name, I.D.

GRIFFITH 101

Location

Owensville PD

AS IV Serial no: 087978
Version no: 532B

TEST RECORD 00784

Temp Date Time ^{s/} 210L

Air Blank:
04/18/24 14:30 .000
Calibration Check:
21 04/18/24 14:30 .103

Subject Name

Subject I.D.

Check 2

Operator Name, I.D.

GRIFFITH 101

Location

Owensville PD

AS IV Serial no: 087978
Version no: 532B

TEST RECORD 00785

Temp Date Time ^{s/} 210L

Air Blank:
04/18/24 14:32 .000
Calibration Check:
22 04/18/24 14:32 .102

Subject Name

Subject I.D.

Check 3

Operator Name, I.D.

GRIFFITH 101

Location

Owensville PD

AS IV Serial no: 087978
Version no: 532B

TEST RECORD 00786

Temp Date Time ^{s/} 210L

VOID: RFI
12 04/18/24 14:35

Subject Name

Subject I.D.

Check - RFI

Operator Name, I.D.

GRIFFITH 101

Location

Owensville PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JONATHAN S. GRIFFITH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/2/2023

NUMBER 230060

EXPIRES 4/2/2025

MO 588-9771 (6-10)

Mike Mason
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielsen
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-16)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **GRIFFITH, JONATHAN**
Permit No **230060**
Date Issued **4/2/2023** Date Expires **4/2/2025**

