



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**RECEIVED**

By Tracy Crews at 10:21 am, Aug 21, 2024

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 096228	NAME OF AGENCY Normandy PD	DATE OF INSPECTION 08/19/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 7700 Natural Bridge Road Normandy, MO 63121		TIME OF INSPECTION 16:13

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories Inc. LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIM. SN SD1615 SIM. NIST EXP DATE 05/21/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  0.099

TEST 2  0.100

TEST 3  0.101

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE  337

PRINT NAME  
Ryan Mack

TYPE II PERMIT NUMBER/EXPIRATION DATE  
240121 05/29/2026

TELEPHONE NUMBER  
(630) 427-7321

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 096228  
Version no: 532B

TEST RECORD 01079

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/19/24 16:13 .000  
Calibration Check:  
23 08/19/24 16:13 .099

Subject Name  
Test #1

Subject I.D.

Maintenance

Operator Name, I.D.

Ryan Mack / 240121

Location

7700 Natural Bridge Rd

Normandy, MO 63121

AS IV Serial no: 096228  
Version no: 532B

TEST RECORD 01080

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/19/24 16:15 .000  
Calibration Check:  
23 08/19/24 16:15 .100

Subject Name

Test #2

Subject I.D.

Maintenance

Operator Name, I.D.

Ryan Mack / 240121

Location

7700 Natural Bridge Rd.

Normandy, MO 63121

AS IV Serial no: 096228  
Version no: 532B

TEST RECORD 01081

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
08/19/24 16:18 .000  
Calibration Check:  
24 08/19/24 16:18 .101

Subject Name

Test #3

Subject I.D.

Maintenance

Operator Name, I.D.

Ryan Mack / 240121

Location

7700 Natural Bridge Rd.

Normandy, MO 63121

AS IV Serial no: 096228  
Version no: 532B

TEST RECORD 01082

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 08/19/24 16:20

Subject Name

RFI Test

Subject I.D.

Maintenance


Operator Name, I.D.

Ryan Mack / 240121

Location

7700 Natural Bridge Rd

Normandy, MO 63121



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L**  $\pm$  3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**RYAN MACK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/29/2024

NUMBER 240121

EXPIRES 5/29/2026

*Mike Maxson*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator MACK, RYAN  
 Permit No 240121  
 Date Issued 5/29/2024 Date Expires 5/29/2026

