



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097417	NAME OF AGENCY Warrenton Police Department	DATE OF INSPECTION 11/20/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 200 West Booneslick Road Warrenton, Missouri 63383		TIME OF INSPECTION 7:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIM. SN MP3585 SIM. NIST EXP DATE 01/30/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .098

TEST 2 ➡ .098

TEST 3 ➡ .097

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

1 DWI Drug Intoxication
 Time was adjusted (was off by approx. 1 hour 45 minutes)

INSPECTING OFFICER

SIGNATURE *Det. James R. Plackemeier*

PRINT NAME
 Det. James R. Plackemeier, DSN 224

TYPE II PERMIT NUMBER/EXPIRATION DATE
 230326 / 12/21/2025

TELEPHONE NUMBER
 (636) 456-3535

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 097417
Version no: 532B

TEST RECORD 00815
s/

Temp Date Time 210L

Air Blank:

11/20/24 07:35 .000

Calibration Check:

21 11/20/24 07:35 .098

Subject Name

Maint Check

Subject I.D.

Test 1

Operator Name, I.D.

Det. Plackemeier, 224

Location

Warrenton PD

AS IV Serial no: 097417
Version no: 532B

TEST RECORD 00816
s/

Temp Date Time 210L

Air Blank:

11/20/24 07:38 .000

Calibration Check:

23 11/20/24 07:38 .098

Subject Name

Maint Check

Subject I.D.

Test 2

Operator Name, I.D.

Det. Plackemeier, 224

Location

Warrenton PD

AS IV Serial no: 097417
Version no: 532B

TEST RECORD 00817
s/

Temp Date Time 210L

Air Blank:

11/20/24 07:41 .000

Calibration Check:

23 11/20/24 07:41 .097

Subject Name

Maint Check

Subject I.D.

Test 3

Operator Name, I.D.

Det. Plackemeier, 224

Location

Warrenton PD

AS IV Serial no: 097417
Version no: 532B

TEST RECORD 00818
s/

Temp Date Time 210L

VOID: RFI

12 11/20/24 07:44

Subject Name

Maint Check

Subject I.D.

Test 4 / RFI

Operator Name, I.D.

Det. Plackemeier, 224

Location

Warrenton PD



Missouri Department of Health and Senior Services

201 Bay Street, Jefferson City, Missouri 65101 Phone: 877.735.4400 Fax: 877.735.4401
201 West Main Street, Hearing and Speech Center, Jefferson City, Missouri 65101

Paula Nickelson
Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP3585 Manufacturer: 3M
Model Number: 12V500
Agency: WARRENTON PD
Agency Address: 211 W. BOONE SLEEK ROAD, WARRENTON, MO 63383

NIST THERMOMETER INFORMATION

Serial Number: 17KMM10696 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 06-27-2023 Date of Expiration: 06-27-2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.06	34.01	0.05

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 1/30/2024
Certification Expiration: 1/31/2025
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: MP3585_1/30/2024

X

DHSS BAP Scientist Approving



GUTH LABORATORIES, INC.

590 NORTH 57th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03072301** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JAMES PLACKEMEIER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

Mike Mason
 DIRECTOR OF STATE PUBLIC HEALTH-LABORATORY

NUMBER 230326

David J. Nielson

EXPIRES 12/21/2025

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PLACKEMEIER, JAMES
Permit No 230326
Date Issued 12/21/2023 **Date Expires** 12/21/2025

