



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>097418</b>	NAME OF AGENCY <b>FENCUS PD</b>	DATE OF INSPECTION <b>08-15-2024</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>222 S PRAIRIE CANT RD FENCUS MO 63735</b>		TIME OF INSPECTION <b>1515</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **21.0°C**

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

STANDARD SUPPLIER **CUTY 495**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°** SIM. SN **501224** SIM. NIST EXP DATE **05-06-2025**

**RECEIVED**  
 By Tracy Crews at 2:54 pm, Aug 16, 2024  
 LOT # **23380** EXP. DATE **10-17-2025**

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <b>.097</b>	TEST 2 • <b>.099</b>	TEST 3 • <b>.100</b>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	<b>0</b>	(0-.04)	<b>0</b>	(.05-.09)	<b>0</b>	(.10-.14)	<b>0</b>	(.15-.19)	<b>0</b>	(OVER .19)	<b>0</b>
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME <b>TRACY CREWS</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>230254 11/14/2025</b>	TELEPHONE NUMBER <b>314-393-8993</b>

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2



**PERMIT**  
**TYPE II**  
**DANNY E. WEST**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/14/2023

NUMBER 230254

EXPIRES 11/14/2025

MO 580-0771 (6-10)

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WEST, DANNY  
Permit No 230254  
Date Issued 11/14/2023 Date Expires 11/14/2025



AS IV Serial no: 097418  
Version no: 532B

TEST RECORD 00186

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/15/24 15:21 .000  
Calibration Check:  
21 08/15/24 15:21 .097

Subject Name

Subject I.D.

Operator Name, I.D.

WEST, DAN  
Location

AS IV Serial no: 097418  
Version no: 532B

TEST RECORD 00187

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/15/24 15:23 .000  
Calibration Check:  
21 08/15/24 15:23 .099

Subject Name

Subject I.D.

Operator Name, I.D.

WEST, DAN  
Location

AS IV Serial no: 097418  
Version no: 532B

TEST RECORD 00188

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/15/24 15:25 .000  
Calibration Check:  
22 08/15/24 15:25 .100

Subject Name

Subject I.D.

Operator Name, I.D.

WEST, DAN  
Location

AS IV Serial no: 097418  
Version no: 532B

TEST RECORD 00189

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 08/15/24 15:26

Subject Name

Subject I.D.

Operator Name, I.D.

WEST, DAN  
Location



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*