



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 099362	NAME OF AGENCY 509 SFS WAFB	DATE OF INSPECTION 28 OCT 2024
LOCATION OF INSTRUMENT (STREET AND CITY) 1031 VANDENBERG AVE, WHITEMAN AFB		TIME OF INSPECTION

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **GUTH** LOT # **22310** EXP. DATE **11 AUG 2024**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34°C** SIM. SN **MP6026** SIM. NIST EXP DATE **26 MAY 25'**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = **.098%**

TEST 2 = **.097%**

TEST 3 = **.097%**

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS **0** (0-.04) **0** (.05-.09) **0** (.10-.14) **0** (.15-.19) **0** (OVER .19) **0**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
BRADY L. GRIMM

TYPE II PERMIT NUMBER/EXPIRATION DATE
240183

TELEPHONE NUMBER
660 687 1459

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 099362
Version no: 532B

TEST RECORD 01897
Temp Date Time 210L s/

Air Blank: 10/28/24 07:18 .000
Subject Test: Auto 25 10/28/24 07:18 .098

Subject Name
BRADY L GRIMM
Subject I.D.

240183 / 29 AUG 2026
Operator Name, I.D.

Location

AS IV Serial no: 099362
Version no: 532B

TEST RECORD 01898
Temp Date Time 210L s/

Air Blank: 10/28/24 07:20 .000
Subject Test: Auto 26 10/28/24 07:20 .097

Subject Name
BRADY L GRIMM
Subject I.D.

240183 / 29 AUG 2026
Operator Name, I.D.

Location

AS IV Serial no: 099362
Version no: 532B

TEST RECORD 01899
Temp Date Time 210L s/

Air Blank: 10/28/24 07:22 .000
Subject Test: Auto 26 10/28/24 07:22 .097

Subject Name
BRADY L GRIMM
Subject I.D.

240183 / 29 AUG 2026
Operator Name, I.D.

Location

AS IV Serial no: 099362
Version no: 532B

TEST RECORD 01900
Temp Date Time 210L s/

UOIH: PJY
12 10/28/24 07:24

Subject Name
BRADY L GRIMM
Subject I.D.

240183 / 29 AUG 2026
Operator Name, I.D.

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22310** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **August 16, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1205%** (w/vol) ethyl alcohol. The expiration date for this lot number is **August 11, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRADY GRIMM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/29/2024

Mike Maxson

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 240183

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/29/2026

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GRIMM, BRADY
 Permit No 240183
 Date Issued 8/29/2024 Date Expires 8/29/2026

