

AS IV
Version:

TEST RECORD 00253

Temp Date Time 210L ^{g/}

Air Blank:
05/28/24 01:27 .000
Calibration Check:
19 05/28/24 01:27 .100

Subject Name

Test #1

Subject I.D.

220276

Operator Name, I.D.

J. Harner

Location

MJC

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00253

Temp Date Time 210L ^{g/}

Air Blank:
05/28/24 01:30 .000
Calibration Check:
21 05/28/24 01:30 .100

Subject Name

Test #3

Subject I.D.

220276

Operator Name, I.D.

J. Harner

Location

MJC

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00252

Temp Date Time 210L ^{g/}

Air Blank:
05/28/24 01:29 .000
Calibration Check:
20 05/28/24 01:29 .100

Subject Name

Test #2

Subject I.D.

220276

Operator Name, I.D.

J. Harner

Location

MJC

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00254

Temp Date Time 210L ^{g/}

VOID: RFI
12 05/28/24 01:33

Subject Name

Test #4

Subject I.D.

220276

Operator Name, I.D.

J. Harner

Location

MJC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 29-Nov-2023

Lot # AG333203 **Model** 108

Exp Date 28-Nov-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: NDIR

Approved for Release: _____
 Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JALYNN HARNER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE 12/21/2022

NUMBER 220276

EXPIRES 12/21/2024

Mike Maszara

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (9-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HARNER, JALYNN
Permit No 220276
Date Issued 12/21/2022 Date Expires 12/21/2024

