



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 07/10/2024 10:30 AM

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>102455</b>	NAME OF AGENCY <b>INDEPENDENCE PD</b>	DATE OF INSPECTION <b>07/03/2024</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>14619 E TROMAN RD, INDEPENDENCE, MO</b>		TIME OF INSPECTION <b>1516</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION                       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **INT-XIMETER**                      LOT # **AG311602**                      EXP. DATE **APRIL 26<sup>th</sup>, 2025**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

**CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <b>.099</b>	TEST 2 • <b>.099</b>	TEST 3 • <b>.099</b>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME <b>BRETT SCHMIDT</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240084                      4/2/2026</b>	TELEPHONE NUMBER <b>816-325-7300</b>

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 102455  
Version no: 532B

TEST RECORD 00318

Temp Date Time %/  
210L

Air Blank:  
07/03/24 15:04 .000  
Calibration Check:  
28 07/03/24 15:04 .099

Subject Name  
TEST 2  
Subject I.D. 240084  
B. SCHMIDT  
Operator Name, I.D.  
14619 E TRUMAN RD  
Location  
MONTHLY MAINTENANCE

AS IV Serial no: 102455  
Version no: 532B

TEST RECORD 00317

Temp Date Time %/  
210L

Air Blank:  
07/03/24 15:01 .000  
Calibration Check:  
28 07/03/24 15:01 .099

Subject Name  
TEST 1  
Subject I.D.  
B. SCHMIDT  
Operator Name, I.D. 240084  
14619 E TRUMAN RD  
Location  
MONTHLY MAINTENANCE

VERSION no: 532B  
TEST RECORD 00319  
Temp Date Time %/  
210L  
Air Blank:  
07/03/24 15:06 .000  
Calibration Check:  
28 07/03/24 15:06 .099  
Subject Name  
TEST 3  
Subject I.D. 240084  
B. SCHMIDT  
Operator Name, I.D.  
14619 E TRUMAN RD  
Location  
MONTHLY MAINTENANCE

AS IV Serial no: 102455  
Version no: 532B  
TEST RECORD 00320  
Temp Date Time %/  
210L  
VOID: RFI  
12 07/03/24 15:09  
Subject Name  
TEST 4  
Subject I.D. 240084  
B. SCHMIDT  
Operator Name, I.D.  
14619 E TRUMAN RD  
Location  
MONTHLY MAINTENANCE



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 26-Apr-2023

Lot # AG311602 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
26-Apr-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 04.26.2023 16:51

Approved for Release:   
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRETT SCHMIDLI**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/2/2024

NUMBER 240084

EXPIRES 4/2/2026

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** SCHMIDLI, BRETT  
**Permit No** 240084  
**Date Issued** 4/2/2024 **Date Expires** 4/2/2026

