



RECEIVED
 By Tracy Crews at 7:31 am, Jul 30, 2024

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102463	NAME OF AGENCY MSHP-E	DATE OF INSPECTION 07/29/2024
-----------------------------	--------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Troop E Headquarters, 4947 HWY 67N, Poplar Bluff, Missouri 63901	TIME OF INSPECTION 8:54 am
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
--	---

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories Inc.</u> LOT # <u>23390</u> EXP. DATE <u>10/17/2025</u>
--

<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.01</u> SIM. SN <u>MP2217</u> SIM. NIST EXP DATE <u>02/28/2025</u>
--

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .099	TEST 2 ← .098	TEST 3 ← .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument operated within DHSS standards.

INSPECTING OFFICER

SIGNATURE SK Tallent	PRINT NAME Shayne K. Talburt
-------------------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230276 11/28/2025	TELEPHONE NUMBER (573) 840-9500
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00324

Temp Date Time 210L

Air Blank: 07/29/24 09:20 .000
Calibration Check: 23 07/29/24 09:20 .099

Subject Name

MAINT CHECK

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT 1110

Location

TROOP E HQ

POPLAR BLUFF, MD

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00323

Temp Date Time 210L

Air Blank: 07/29/24 09:15 .000
Calibration Check: 22 07/29/24 09:15 .098

Subject Name

MAINT CHECK

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT

Location

TROOP E HQ

POPLAR BLUFF, MD

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00322

Temp Date Time 210L

Air Blank: 07/29/24 09:12 .000
Calibration Check: 21 07/29/24 09:12 .099

Subject Name

MAINT CHECK

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT 1110

Location

TROOP E

POPLAR BLUFF, MD

POPLAR BLUFF, MD
TROOP E HQ
Location
S.K. TALBUNT 1110
Operator Name, I.D.

MAINT CHECK
Subject Name
12 07/29/24 09:25
Temp Date Time 210L
TEST RECORD 00325

AS IV Serial no: 102463
Version no: 532B



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
SHAYNE K. TALBURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/28/2023

NUMBER 230276

EXPIRES 11/28/2025

MO 580-0771 (6-10)

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R5-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TALBURT, SHAYNE

Permit No 230276

Date Issued 11/28/2023 Date Expires 11/28/2025

