

RECEIVED

Get Original

By Tracy Crews at 9:20 am, Aug 26, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102467	NAME OF AGENCY Camden County SO	DATE OF INSPECTION 08/16/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 1 Court Circle NW, Camdenton, Mo 65020		TIME OF INSPECTION 0530

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 23390 EXP. DATE 10/17/25

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN SD2758 SIM. NIST EXP DATE 01/29/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .104	TEST 2 .103	TEST 3 .103
--------------------	--------------------	--------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(.0-.04) 0	(.05-.09) 1	(.10-.14) 16	(.15-.19) 5	(OVER .19) 4
----------	-------------------	--------------------	---------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE ▶	PRINT NAME
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER ()

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 102467
Version no: 532B

TEST RECORD 00551

Temp Date Time ^{a/} 210L

Air Blank:
08/16/24 05:30 .000
Calibration Check:
20 08/16/24 05:30 .104

Subject Name

Test 1

Subject I.D.

AS IV Serial no: 102467
Version no: 532B

TEST RECORD 00552

Temp Date Time ^{a/} 210L

Air Blank:
08/16/24 05:33 .000
Calibration Check:
21 08/16/24 05:33 .103

Subject Name

Test 2

Subject I.D.

AS IV Serial no: 102467
Version no: 532B

TEST RECORD 00553

Temp Date Time ^{a/} 210L

Air Blank:
08/16/24 05:35 .000
Calibration Check:
21 08/16/24 05:35 .103

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Bryce Easley 23094

Location

1 Court Cir NW

Camdenton, Mo 65020

Operator Name, I.D.

Bryce Easley 23094

Location

1 Court Cir NW

Camdenton, Mo 65020

Operator Name, I.D.

Bryce Easley 23094

Location

1 Court Cir NW

Camdenton, Mo 65020

AS IV Serial no: 102467
Version no: 532B

TEST RECORD 00554

Temp Date Time ^{a/} 210L

VOID: RFI
12 08/16/24 05:37

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Bryce Easley 23094

Location

1 Court Cir NW

Camdenton, Mo 65020



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03072301** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRYCE EASLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/23/2023

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230094

David J. Nickelson

EXPIRES 5/23/2025

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator EASLEY, BRYCE
Permit No 230094
Date Issued 5/23/2023 **Date Expires** 5/23/2025

