



AS IV Serial no: 104642  
Version no: 532B

TEST RECORD 01234

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
12/04/24 23:53 .000  
Calibration Check:  
21 12/04/24 23:53 .080

Subject Name  
TEST 1  
Subject I.D.

Operator Name, I.D.  
SINNOGRAK 334  
Location

AS IV Serial no: 104642  
Version no: 532B

TEST RECORD 01235

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
12/04/24 23:55 .000  
Calibration Check:  
22 12/04/24 23:55 .079

Subject Name  
TEST 2  
Subject I.D.

Operator Name, I.D.  
SINNOGRAK 334  
Location

AS IV Serial no: 104642  
Version no: 532B

TEST RECORD 01236

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
12/04/24 23:57 .000  
Calibration Check:  
23 12/04/24 23:57 .079

Subject Name  
TEST 3  
Subject I.D.

Operator Name, I.D.  
SINNOGRAK 334  
Location

AS IV Serial no: 104642  
Version no: 532B

TEST RECORD 01237

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 12/04/24 23:59

Subject Name  
RFI  
Subject I.D.

Operator Name, I.D.  
  
Location

AS IV Serial no: 104642  
Version no: 532B

TEST RECORD 01238

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
12/05/24 00:01 .000  
Subject Test: Auto  
23 12/05/24 00:01 .000

Subject Name  
SBAR RFI  
Subject I.D.

Operator Name, I.D.  
  
Location



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 18-Apr-2023

**Lot #** AG310305 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
13-Apr-2025	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control  
Reason:Dry gas standard certification of analysis  
Location:Airgas USA LLC (Lab)  
Date:04.19.2023 17:13

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**CHRIS SINNOKRAK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/8/2024

NUMBER 240006

EXPIRES 1/8/2026

*Mike Massum*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Dave J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** SINNOKRAK, CHRIS  
**Permit No** 240006  
**Date Issued** 1/8/2024 **Date Expires** 1/8/2026

