

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01895

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/10/24 03:50 .000
Calibration Check:
23 06/10/24 03:50 .098

Subject Name
TEST #1

Subject I.D.
N/A

Operator Name, I.D.

M. Nelson #220186
Location

Clinton P.D.
101 E. Ohio
Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01896

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/10/24 03:52 .000
Calibration Check:
24 06/10/24 03:52 .098

Subject Name
TEST #2

Subject I.D.
N/A

Operator Name, I.D.

M. Nelson #220186
Location

Clinton P.D.
101 E. Ohio
Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01897

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/10/24 03:53 .000
Calibration Check:
25 06/10/24 03:53 .098

Subject Name
TEST #3

Subject I.D.
N/A

Operator Name, I.D.

M. Nelson #220186
Location

Clinton P.D.
101 E. Ohio
Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01898

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 06/10/24 03:55

Subject Name
RFI

Subject I.D.
N/A

Operator Name, I.D.

M. Nelson #220186
Location

Clinton P.D.
101 E. Ohio
Clinton, MO 64735

AS IV Serial no: 105448
Version no: 532B

TEST RECORD 01899

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
06/10/24 03:56 .000
Subject Test: Auto
26 06/10/24 03:56 .000

Subject Name
SELF TEST

Subject I.D.
N/A

Operator Name, I.D.

M. Nelson #220186
Location

Clinton P.D.
101 E. Ohio
Clinton, MO 64735



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23390 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1207% (w/vol) ethyl alcohol. The expiration date for this lot number is October 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

MICHAEL S. NELSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2022

NUMBER 220186

EXPIRES 7/22/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RIS-10)

MO 680-0771 (8-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **NELSON, MICHAEL**
 Permit No **220186**
 Date Issued **7/22/2022** Date Expires **7/22/2024**

