

RECEIVED

By Tracy Crews at 7:34 am, Oct 02, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107977	NAME OF AGENCY Excelsior Springs Police Department	DATE OF INSPECTION 10/01/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 301 S Main St, Excelsior Springs, MO 64024		TIME OF INSPECTION 0850

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 C SIM. SN MP3564 SIM. NIST EXP DATE 05/16/2024

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100 TEST 2 ← .100 TEST 3 ← .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 2 (0-.04) 0 (.05-.09) 0 (.10-.14) 1 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Brian K Kennedy

TYPE II PERMIT NUMBER/EXPIRATION DATE

230039 03/09/2025

TELEPHONE NUMBER

816-629-7108

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 107977
Version no: 532B

TEST RECORD 01453

Temp Date Time 210L
s/

VOID: RFI
12 10/01/24 09:00

Subject Name
SIMULATOR
Subject I.D.

Operator Name, I.D.
KENNEDY 911
Location
SIMULATOR

AS IV Serial no: 107977
Version no: 532B

TEST RECORD 01452

Temp Date Time 210L
s/

Air Blank:
10/01/24 08:59 .000
Calibration Check:
22 10/01/24 08:59 .100

Subject Name
SIMULATOR
Subject I.D.

Operator Name, I.D.
KENNEDY 911
Location
ESPD

AS IV Serial no: 107977
Version no: 532B

TEST RECORD 01451

Temp Date Time 210L
s/

Air Blank:
10/01/24 08:57 .000
Calibration Check:
22 10/01/24 08:57 .100

Subject Name
SIMULATOR
Subject I.D.

Operator Name, I.D.
KENNEDY 911
Location
ESAD

AS IV Serial no: 107977
Version no: 532B

TEST RECORD 01450

Temp Date Time 210L
s/

Air Blank:
10/01/24 08:56 .000
Calibration Check:
21 10/01/24 08:56 .100

Subject Name
SIMULATOR
Subject I.D.

Operator Name, I.D.
KENNEDY 911
Location
ESPD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

BRIAN K. KENNEDY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2023

NUMBER 230039

EXPIRES 3/9/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KENNEDY, BRIAN
Permit No 230039
Date Issued 3/9/2023 **Date Expires** 3/9/2025

