



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107984	NAME OF AGENCY MSHP	DATE OF INSPECTION 12/04/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 1702 E. Laharpe Street, Kirksville, MO 63501		TIME OF INSPECTION 9:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 23390 EXP. DATE 10/17/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN MP2519 SIM. NIST EXP DATE 04/25/2025
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .095	TEST 2 ➔ .096	TEST 3 ➔ .095
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <i>[Signature]</i>	PRINT NAME Tpr. C. J. Snyder, #491
TYPE II PERMIT NUMBER/EXPIRATION DATE 240208 09/10/2026	TELEPHONE NUMBER (660) 385-2132

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

DEC 04 2024

AS IV Serial no: 107984
Version no: 532B

TEST RECORD 00870

Temp Date Time 210L
s/

Air Blank: 12/04/24 09:35 .000
Calibration Check: 22 12/04/24 09:35 .095

Subject Name
Test # 1
Subject I.D.

Operator Name, I.D.

Tr. C.J. Snyder 240209
Location

1702 G. Lohrpe St.

Kirksville, MO

Cody G. Snyder
441

AS IV Serial no: 107984
Version no: 532B

TEST RECORD 00871

Temp Date Time 210L
s/

Air Blank: 12/04/24 09:38 .000
Calibration Check: 22 12/04/24 09:38 .096

Subject Name
Test # 2
Subject I.D.

Operator Name, I.D.

Tr. C.J. Snyder 240209
Location

1702 G. Lohrpe St.

Kirksville, MO

Cody G. Snyder
441

AS IV Serial no: 107984
Version no: 532B

TEST RECORD 00872

Temp Date Time 210L
s/

Air Blank: 12/04/24 09:40 .000
Calibration Check: 22 12/04/24 09:40 .095

Subject Name
Test # 3
Subject I.D.

Operator Name, I.D.

Tr. C.J. Snyder 240208
Location

1702 G. Lohrpe St.

Kirksville, MO

Cody G. Snyder
441

AS IV Serial no: 107984
Version no: 532B

TEST RECORD 00873

Temp Date Time 210L
s/

VOID: RFI
12 12/04/24 09:41

Subject Name
RFI
Subject I.D.

Operator Name, I.D.

Tr. C.J. Snyder 240208
Location

1702 G. Lohrpe St.

Kirksville, MO

Cody G. Snyder
441



GUTH LABORATORIES, INC.

390 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23390 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XI, S/N: 610N9030209, and found to contain 0.1207% (w/vol) ethyl alcohol. The expiration date for this lot number is October 17, 2025 at 11:59 P.M.

When used in a calibrated Simulator, operating at 34°C +/- 0.2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Certified Reference Standard lot number EN03072301 whose values are traceable to NIST

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
 P O Box 570, Jefferson City, MO 65102 0570 Phone 573 751 6400 FAX 573 751 6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800 735 7166 VOICE 1 800 735 2364

Paula Nickelson
 Acting Director

Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2519 Manufacturer: Guth
 Model Number: 12V500
 Agency: MSHP (GHQ)
 Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

NIST THERMOMETER INFORMATION

Serial Number: 307715 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 10/4 2023 Date of Expiration: 10/4 2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
33.99	34.02	.06

The combined uncertainty is calculated with a k_c 2 value

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 4/25/2024
 Certification Expiration: 4/25/2025
 Simulator testing technician: J. CLEVELAND

Notes on Condition: none
 Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
 Certification No: MP2519 4252024

X *Brianna Medrano*

DHSS BAP Scientist Approving

Simulator Calibration Certification
 Issued by Lab Manager, DHSS BAP
 Revision Date: 06/25/2022

Breath Alcohol Program
 1903 Northwood Drive, Suite 4
 Poplar Bluff, MO 63901

DHSS BAP Document 3.6A
 Revision 2
 Page 1 of 1



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CODY SNYDER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/10/2024

NUMBER 240208

EXPIRES 9/10/2026

Adam H. ...

DIRECTOR STATE PUBLIC HEALTH LABORATORY

David J. ...

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator SNYDER, CODY
 Permit No 240208
 Date Issued 9/10/2024 Date Expires 9/10/2026

