



**RECEIVED**

By Tracy Crews at 9:35 am, Aug 09, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111318	NAME OF AGENCY GREENE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 08/02/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 5100 W DIVISION ST SPRINGFIELD MO 65802		TIME OF INSPECTION 10:06 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG331103 EXP. DATE 11/07/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → 0.101 g/210L	TEST 2 → 0.101 g/210L	TEST 3 → 0.100 g/210L
-----------------------	-----------------------	-----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME JEREMY GRISHAM
TYPE II PERMIT NUMBER/EXPIRATION DATE TYPE II PERMIT #240077 EXPIRATION 04/02/2026	TELEPHONE NUMBER (417) 868-4040

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

ASTV Serial no: 111318  
Version no: 532B

TEST RECORD 00729

Temp Date Time 9/21/01

Air Blank:  
08/02/24 10:08 000  
Calibration Check:  
21 08/02/24 10:08 100

Subject Name

ACC TEST  
Subject T.D.

TEST 1  
Operator Name, T.D.

GRISHAM 240077 4/2/26  
Location

GCSO HQ

5700 W Division St

ASTV Serial no: 111318  
Version no: 532B

TEST RECORD 00730

Temp Date Time 9/21/01

Air Blank:  
08/02/24 10:08 000  
Calibration Check:  
21 08/02/24 10:08 100

Subject Name

ACC TEST  
Subject T.D.

TEST 2  
Operator Name, T.D.

GRISHAM 240077 4/2/26  
Location

GCSO HQ

5700 W Division St

ASTV Serial no: 111318  
Version no: 532B

TEST RECORD 00731

Temp Date Time 9/21/01

Air Blank:  
08/02/24 10:10 000  
Calibration Check:  
21 08/02/24 10:10 100

Subject Name

ACC TEST  
Subject T.D.

TEST 3  
Operator Name, T.D.

GRISHAM 240077 4/2/26  
Location

GCSO HQ

5700 W Division

ASTV Serial no: 111318  
Version no: 532B

TEST RECORD 00732

Temp Date Time 9/21/01

VOID: RFI  
12 08/02/24 10:11

Subject Name

RFI TEST  
Subject T.D.

TEST #4  
Operator Name, T.D.

GRISHAM 240077 4/2/26  
Location

GCSO HQ

5700 W Division

ASTV Serial no: 111318  
Version no: 532B

TEST RECORD 00733

Temp Date Time 9/21/01

Air Blank:  
08/02/24 10:13 000  
Subject Test: Auto  
21 08/02/24 10:13 000

Subject Name

SOBER/CONTIN  
Subject T.D.

TEST #5  
Operator Name, T.D.

GRISHAM 240077 4/2/26  
Location

GCSO - 5700 W Division



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 8-Nov-2023

**Lot #** AG331103 **Model** 108

<b>Exp Date</b> 7-Nov-2025	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
-------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:11.09.2023 19:42

Approved for Release: \_\_\_\_\_

Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JEREMY GRISHAM**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/2/2024

NUMBER 240077

EXPIRES 4/2/2026

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator GRISHAM, JEREMY  
Permit No 240077  
Date Issued 4/2/2024 Date Expires 4/2/2026

