



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 2:22 pm, Nov 20, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111331	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 11/15/2024
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 2920 N. Shamrock Rd, Jefferson City	TIME OF INSPECTION 10:25 am
---	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIM. SN MP2127 SIM. NIST EXP DATE 10/03/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100

TEST 2 .100

TEST 3 .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	1	(.05-.09)	4	(.10-.14)	3	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Breath manifold replaced and Instrument calibrated.

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Jimmy L. Cleveland

TYPE OF PERMIT NUMBER/EXPIRATION DATE
230082 / 050325

TELEPHONE NUMBER
573-30-751-4722

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111331
Version no: 532B

TEST RECORD 01226

Temp Date Time ^{s/} 210L

Air Blank:
11/15/24 10:27 .000
Calibration Check:
24 11/15/24 10:27 .100

Subject Name

CALIBRATION

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111331
Version no: 532B

TEST RECORD 01227

Temp Date Time ^{s/} 210L

Air Blank:
11/15/24 10:35 .000
Calibration Check:
24 11/15/24 10:35 .100

Subject Name

Acc 1

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111331
Version no: 532B

TEST RECORD 01228

Temp Date Time ^{s/} 210L

Air Blank:
11/15/24 10:40 .000
Calibration Check:
25 11/15/24 10:40 .100

Subject Name

Acc 2

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111331
Version no: 532B

TEST RECORD 01229

Temp Date Time ^{s/} 210L

Air Blank:
11/15/24 10:47 .000
Calibration Check:
24 11/15/24 10:47 .100

Subject Name

Acc 3

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 111331
Version no: 532B

TEST RECORD 01230

Temp Date Time ^{s/} 210L

VOID: RFI
12 11/15/24 10:48

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

JIMMY L. CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and re and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sec 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/3/2023

NUMBER 230082

EXPIRES 5/3/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CLEVELAND, JIMMY
 Permit No 230082
 Date Issued 5/3/2023 Date Expires 5/3/2025

