

AS IV Serial no: 111332
Version no: 532B

TEST RECORD 01238

Temp Date Time ^{a/} 210L

Air Blank:
06/03/24 13:02 .000
Calibration Check:
21 06/03/24 13:02 .101

Subject Name

Test

Subject I.D.

Test 1

Operator Name, I.D.

William Housley 23068

Location

517 Oneida St.

Seneca Mo 64865

AS IV Serial no: 111332
Version no: 532B

TEST RECORD 01239

Temp Date Time ^{a/} 210L

Air Blank:
06/03/24 13:03 .000
Calibration Check:
21 06/03/24 13:03 .101

Subject Name

Test

Subject I.D.

Test 2

Operator Name, I.D.

William Housley 23068

Location

517 Oneida St.

Seneca Mo 64865

AS IV Serial no: 111332
Version no: 532B

TEST RECORD 01240

Temp Date Time ^{a/} 210L

Air Blank:
06/03/24 13:05 .000
Calibration Check:
22 06/03/24 13:05 .100

Subject Name

Test

Subject I.D.

Test 3

Operator Name, I.D.

William Housley 23068

Location

517 Oneida St.

Seneca Mo 64865

AS IV Serial no: 111332
Version no: 532B

TEST RECORD 01241

Temp Date Time ^{a/} 210L

VOID: RFI
12 06/03/24 13:06

Subject Name

Test

Subject I.D.

RFI Test

Operator Name, I.D.

William Housley 23068

Location

517 Oneida St.

Seneca Mo 64865



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
WILLIAM J. HOUSLEY II

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/7/2023

NUMBER 230168

EXPIRES 8/7/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Rickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HOUSLEY II, WILLIAM
Permit No 230168
Date Issued 8/7/2023 **Date Expires** 8/7/2025





GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.