

RECEIVED

By Tracy Crews at 9:13 am, Jun 10, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111335 NAME OF AGENCY St. Louis County Police Department DATE OF INSPECTION 06/05/2024

LOCATION OF INSTRUMENT (STREET AND CITY) 14301 South Outer Fourty, Chesterfield Missouri TIME OF INSPECTION 7:50 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

[X] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

[X] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

[X] PRINTER WORKING PROPERLY

[X] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

[] SIMULATOR SOLUTION [X] COMPRESSED ETHANOL-GAS MIXTURE

[X] STANDARD SUPPLIER INTOXIMETERS LOT # AG321505 EXP. DATE 08/03/2025

[] SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

[X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [X] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
[] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
[] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .096

TEST 2 .096

TEST 3 .095

[X] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE [Handwritten Signature]

PRINT NAME Austin Wesolich DSN 4971

TYPE II PERMIT NUMBER/EXPIRATION DATE 230248 / 11/01/2025

TELEPHONE NUMBER (636) 628-6079

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

Temp Date Time ^{9/} 210L

Air Blank:
06/05/24 07:50 .000
Calibration Check:
23 06/05/24 07:50 .096

Subject Name

TEST 1

Subject I.D.

TEST 1

Operator Name, I.D.

WESOLICH (230248)

Location

MODOT TML

AS IV Serial no: 111335
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00607

Temp Date Time ^{9/} 210L

Air Blank:
06/05/24 07:52 .000
Calibration Check:
23 06/05/24 07:52 .096

Subject Name

TEST 2

Subject I.D.

TEST 2

Operator Name, I.D.

WESOLICH (230248)

Location

MODOT TML

AS IV Serial no: 111335
Version no: 532B

TEST RECORD 00608

Temp Date Time ^{9/} 210L

Air Blank:
06/05/24 07:58 .000
Calibration Check:
23 06/05/24 07:58 .095

Subject Name

TEST 3

Subject I.D.

TEST 3

Operator Name, I.D.

WESOLICH (230248)

Location

MODOT TML

AS IV Serial no: 111335
Version no: 532B

TEST RECORD 00609

Temp Date Time ^{9/} 210L

VOID: RFI
12 06/05/24 08:07

Subject Name

RFI!

Subject I.D.

RFI!

Operator Name, I.D.

WESOLICH (230248)

Location

MODOT TML



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 3-Aug-2023

Lot # AG321505 **Model** 108

| | | | |
|-------------------------------|-------------------------|---|---|
| Exp Date 3-Aug-2025 | Cyl. Type 108 | Component Ethanol Nitrogen | Certified Concentration 0.100 ± 2% BrAC (272 ppm) |
|-------------------------------|-------------------------|---|---|

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010562 | 104.2 ppm |
| EB0010561 | 103.7 ppm | EB0010579 | 52.94 ppm |
| EB0010681 | 52.22 ppm | | |


| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481 | 800.0 ppm | CC727493 | 390.0 ppm |
| CC727496 | 253.0 ppm | CC727498 | 150.0 ppm |

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 08.03.2023 17:58

Approved for Release: _____
 Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

AUSTIN WESOLICH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 11/1/2023

NUMBER: 230248

EXPIRES: 11/1/2025

Mike Mazzamuni

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-590-0774 (8-10)

LAB-4 (R6/10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WESOLICH, AUSTIN
 Permit No 230248
 Date Issued 11/1/2023 Date Expires 11/1/2025

