

AS IV Serial no: 111631
Version no: 532B

TEST RECORD 00263

Temp Date Time ^{9/} 210L

Air Blank:
11/08/24 09:14 .000
Calibration Check:
18 11/08/24 09:14 .104

Subject Name

Test #1

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 111631
Version no: 532B

TEST RECORD 00264

Temp Date Time ^{9/} 210L

Air Blank:
11/08/24 09:16 .000
Calibration Check:
19 11/08/24 09:16 .103

Subject Name

Test #2

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 111631
Version no: 532B

TEST RECORD 00265

Temp Date Time ^{9/} 210L

Air Blank:
11/08/24 09:19 .000
Calibration Check:
21 11/08/24 09:19 .101

Subject Name

Test #3

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 111631
Version no: 532B

TEST RECORD 00266

Temp Date Time ^{9/} 210L

VOID: RFI
12 11/08/24 09:20

Subject Name

RFI Test

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 18-Mar-2024

Lot # AG407603 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
16-Mar-2026	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 03.22.2024 07:48

Approved for Release: _____
 Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
THOMAS F. BUTKOVICH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/5/2023

NUMBER 230113

EXPIRES 6/5/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula F. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BUTKOVICH, THOMAS
Permit No 230113
Date issued 6/5/2023 Date Expires 6/5/2025

