



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>111642</u>	NAME OF AGENCY <u>Campbell Police Department</u>	DATE OF INSPECTION <u>6-4-2024</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>204 Grand Ave Campbell, Mo 63933</u>		TIME OF INSPECTION <u>05:39</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 23390 EXP. DATE 10-17-2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34°C SIM. SN SD1743 SIM. NIST EXP DATE 12-19-2025

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \Rightarrow <u>.096</u>	TEST 2 \Rightarrow <u>.098</u>	TEST 3 \Rightarrow <u>.097</u>
----------------------------------	----------------------------------	----------------------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<u>0</u>	(0-.04)	<u>0</u>	(.05-.09)	<u>0</u>	(.10-.14)	<u>0</u>	(.15-.19)	<u>0</u>	(OVER .19)	<u>0</u>
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

Jerry D. Hargraves

PRINT NAME

Jerry D. Hargraves

TYPE II PERMIT NUMBER/EXPIRATION DATE

230319 / 12-21-2025

TELEPHONE NUMBER

(573) 217-9122

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111642
Version no: 532B

TEST RECORD 00572

Temp Date Time ^{9/} 210L

Air Blank:
06/04/24 05:43 .000
Calibration Check:
19 06/04/24 05:43 .096

Subject Name

test #1

Subject I.D.

Jerry Hargraves 230319
Operator Name, I.D.

Campbell Police Dept
Location

Booking room

AS IV Serial no: 111642
Version no: 532B

TEST RECORD 00573

Temp Date Time ^{9/} 210L

Air Blank:
06/04/24 05:44 .000
Subject Test: Man
20 06/04/24 05:44 .098

Subject Name

test #2

Subject I.D.

Jerry Hargraves 230319
Operator Name, I.D.

Campbell Police Dept.
Location

Booking room

AS IV Serial no: 111642
Version no: 532B

TEST RECORD 00574

Temp Date Time ^{9/} 210L

Air Blank:
06/04/24 05:46 .000
Subject Test: Man
21 06/04/24 05:46 .097

Subject Name

test #3

Subject I.D.

Jerry Hargraves 230319
Operator Name, I.D.

Campbell Police Dept
Location

Booking room

AS IV Serial no: 111642
Version no: 532B

TEST RECORD 00575

Temp Date Time ^{9/} 210L

VOID: RFI
12 06/04/24 05:47

Subject Name

test #4 RFI test

Subject I.D.

Jerry Hargraves 230319
Operator Name, I.D.

Campbell Police Dept
Location

Booking room



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Paula Nickelson
 Acting Director

Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD1743 **Manufacturer:** Guth
Model Number: 10-4D
Agency: CAMPBELL PD
Agency Address: 204 W GRAND, CAMPBELL, MO 63933

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 10/27/2023 **Date of Expiration:** 10/27/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.03	.06

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 12/19/2023
Certification Expiration: 12/19/2024
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none
Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: SD1743_12192023

X *Brianna Medrano*

DHSS BAP Scientist Approving



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JERRY HARGRAVES

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

Mike Masman
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230319

Paula J. Nielson
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 12/21/2025

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HARGRAVES, JERRY
Permit No 230319
Date Issued 12/21/2023 **Date Expires** 12/21/2025

