



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111657	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 11/21/2024
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 27821 RT-7 S, Harrisonville, MO 64701	TIME OF INSPECTION 04:50 p.m.
---	----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 23390 EXP. DATE 10/17/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.98 SIM. SN MP2216 SIM. NIST EXP DATE 11/28/2024

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .104	TEST 3 .103
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
------------	-----------	-------------	-------------	-------------	--------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Clock adjusted 1 hour back for daylight savings.

INSPECTING OFFICER

SIGNATURE #1340	PRINT NAME Cpl. W.J. Luzenko #1340
TYPE II PERMIT NUMBER/EXPIRATION DATE 240141 / 6/21/2026	TELEPHONE NUMBER (816) 622-0800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00629

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/21/24 18:46 .000
Calibration Check:
21 11/21/24 18:46 .103

Subject Name

Maint Test #1

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (240141)

Location

Cass County Zone Office

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00630

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/21/24 18:47 .000
Calibration Check:
22 11/21/24 18:47 .104

Subject Name

Maint Test #2

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (240141)

Location

Cass County Zone Office

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00631

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
11/21/24 18:49 .000
Calibration Check:
22 11/21/24 18:49 .103

Subject Name

Maint Test #3

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (240141)

Location

Cass County Zone Office

AS IV Serial no: 111657
Version no: 532B

TEST RECORD 00632

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 11/21/24 18:50

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Cpl. W.J. Wzenko (240141)

Location

Cass County Zone Office



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

WALTER J. LUZENKO

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/21/2024

NUMBER 240141

EXPIRES 6/21/2026

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680 0771 (6-10)

LAB 4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LUZENKO, WALTER
Permit No 240141
Date Issued 6/21/2024 Date Expires 6/21/2026

