



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|                             |  |                                  |
|-----------------------------|--|----------------------------------|
| ALCO SENSOR IV SN<br>111667 | NAME OF AGENCY<br>Platte County Sheriff's Office | DATE OF INSPECTION<br>10/03/2024 |
|-----------------------------|--|----------------------------------|

|   |                                |
|---|--------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>11724 NW PLAZA CIR, KANSAS CITY | TIME OF INSPECTION<br>10:59 pm |
|---|--------------------------------|

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

|   |  |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION                                     | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters, Inc.</u> | LOT # <u>AG309501</u> EXP. DATE <u>04/05/2025</u>                  |
| <input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)                   | SIM. SN _____ SIM. NIST EXP DATE _____                             |

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

|               |               |               |
|---------------|---------------|---------------|
| TEST 1  0.096 | TEST 2  0.096 | TEST 3  0.096 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |         |           |           |           |            |
|----------|---------|-----------|-----------|-----------|------------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | (OVER .19) |
|----------|---------|-----------|-----------|-----------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

|  |                                    |
|--|------------------------------------|
| SIGNATURE<br>  | PRINT NAME<br>Daniel Gearhart      |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>240092 04/12/2026 | TELEPHONE NUMBER<br>(816) 858-3521 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111667  
Version no: 532B

TEST RECORD 01053

Temp Date Time <sup>g/</sup>210L

Air Blank:  
10/03/24 22:59 .000  
Calibration Check:  
24 10/03/24 22:59 .096

Subject Name

MAINTENANCE

Subject I.D.

TEST #1

Operator Name, I.D.

GEARHART 240092

Location

11724 NW PLAZA CER

KANSAS CITY, MO

AS IV Serial no: 111667  
Version no: 532B

TEST RECORD 01054

Temp Date Time <sup>g/</sup>210L

Air Blank:  
10/03/24 23:00 .000  
Calibration Check:  
24 10/03/24 23:00 .096

Subject Name

MAINTENANCE

Subject I.D.

TEST #2

Operator Name, I.D.

GEARHART 240092

Location

11724 NW PLAZA CER

KANSAS CITY, MO

AS IV Serial no: 111667  
Version no: 532B

TEST RECORD 01055

Temp Date Time <sup>g/</sup>210L

Air Blank:  
10/03/24 23:02 .000  
Calibration Check:  
25 10/03/24 23:02 .096

Subject Name

MAINTENANCE

Subject I.D.

TEST #3

Operator Name, I.D.

GEARHART 240092

Location

11724 NW PLAZA CER

KANSAS CITY, MO

AS IV Serial no: 111667  
Version no: 532B

TEST RECORD 01056

Temp Date Time <sup>g/</sup>210L

VOID: RFI  
12 10/03/24 23:03

Subject Name

MAINTENANCE

Subject I.D.

TEST #4

Operator Name, I.D.

GEARHART 240092

Location

11724 NW PLAZA CER

KANSAS CITY, MO



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Test Date: 5-Apr-2023

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Lot #** AG309501 **Model** 108

|                               |                         |   |   |
|-------------------------------|-------------------------|---|---|
| <b>Exp Date</b><br>5-Apr-2025 | <b>Cyl. Type</b><br>108 | <b>Component</b><br>Ethanol<br>Nitrogen | <b>Certified Concentration</b><br>0.100 ± 2% BrAC (272 ppm) |
|-------------------------------|-------------------------|---|---|

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581      | 391.8 ppm     | EB0010603      | 392.5 ppm     |
| EB0010570      | 259.8 ppm     | EB0010559      | 258.9 ppm     |
| EB0010285      | 209.0 ppm     | EB0010562      | 104.2 ppm     |
| EB0010561      | 103.7 ppm     | EB0010579      | 52.94 ppm     |
| EB0010681      | 52.22 ppm     |                |               |

  

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481       | 800.0 ppm     | CC727493       | 390.0 ppm     |
| CC727496       | 253.0 ppm     | CC727498       | 150.0 ppm     |

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:04.05.2023 17:34

Approved for Release:                     *Rod Marsala*                      
 Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**DANIEL GEARHART**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/12/2024

NUMBER 240092

EXPIRES 4/12/2026

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula F. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri*

Operator GEARHART, DANIEL  
Permit No 240092  
Date Issued 4/12/2024 Date Expires 4/12/2026

