

AS IV Serial no: 111673
Version no: 532B

TEST RECORD 60511

Temp Date Time 21BL

Air Blank
06/25/74 12:24 .000
Calibration Check
21 06/25/74 12:24 .000

Substrate Name

TEST 1

Substrate No.

TEST 1

Operator Name: J.H.

WESOLICH (230248)

Location

MODOT TRC

AS IV Serial no: 111673
Version no: 532B

TEST RECORD 60512

Temp Date Time 21BL

Air Blank
06/27/74 12:26 .000
Calibration Check
21 06/27/74 12:26 .000

Substrate Name

TEST 2

Substrate No.

TEST 2

Operator Name: J.H.

WESOLICH (230248)

Location

MODOT TRC

AS IV Serial no: 111673
Version no: 532B

TEST RECORD 60513

Temp Date Time 21BL

Air Blank
06/27/74 12:27 .000
Calibration Check
21 06/27/74 12:27 .000

Substrate Name

TEST 3

Substrate No.

TEST 3

Operator Name: J.H.

WESOLICH (230248)

Location

MODOT TRC

AS IV Serial no: 111673
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 60514

Temp Date Time 21BL

Air Blank
06/27/74 12:27

Substrate Name

RFI 1

Substrate No.

RFI 1

Operator Name: J.H.

WESOLICH (230248)

Location

MODOT TRC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 3-Aug-2023

Lot # AG321505 **Model** 108

Exp Date 3-Aug-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
-------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm


Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:08.03.2023 17:58

Approved for Release: _____

Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

AUSTIN WESOLICH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.021, RSMo and 308.111 through 308.119 RSMo.

DATE 11/1/2023

NUMBER 230248

EXPIRES 11/1/2025

Mike Magallon

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO:5809771(6-10)

LAB: (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WESOLICH, AUSTIN
Permit No 230248
Date Issued 11/1/2023 Date Expires 11/1/2025

