



By Tracy Crews at 12:11 pm, Aug 02, 2024

RECEIVED

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111749	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 08/02/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 504 S.E. Blue Parkway Lee's Summit, MO 64063		TIME OF INSPECTION 9:10 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABORATORIES INC LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2206 SIM. NIST EXP DATE 11/27/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .099	TEST 3 ← .098
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Benjamin Owens
TYPE II PERMIT NUMBER/EXPIRATION DATE Permit # 230048/ Expiration Date 03/27/2025	TELEPHONE NUMBER (816) 622-0800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office

By Tracy Crews at 12:11 pm, Aug 02, 2024

RECEIVED

AS IV Serial no: 111749
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00901

Temp Date Time 210L
9/

Air Blank: 08/02/24 09:10 .000
Calibration Check: 24 08/02/24 09:10 .100

Subject Name

Test #1

Subject I.D.

Cpl. B. Owens 1423

Operator Name, I.D.

Permit 230048

Location

504 SE Blue Parkway

Leas Summit No

AS IV Serial no: 111749
Version no: 532B

TEST RECORD 00902

Temp Date Time 210L
9/

Air Blank: 08/02/24 09:13 .000
Calibration Check: 25 08/02/24 09:13 .099

Subject Name

Test #2

Subject I.D.

Cpl. B. Owens

Operator Name, I.D.

Permit 230048

Location

504 SE Blue Parkway

Leas Summit No

AS IV Serial no: 111749
Version no: 532B

TEST RECORD 00903

Temp Date Time 210L
9/

Air Blank: 08/02/24 09:14 .000
Calibration Check: 26 08/02/24 09:14 .098

Subject Name

Test #3

Subject I.D.

Cpl. B. Owens

Operator Name, I.D.

Permit 230048

Location

504 SE Blue Parkway

Leas Summit No

AS IV Serial no: 111749
Version no: 532B

TEST RECORD 00904

Temp Date Time 210L
9/

VOID: RFI
12 08/02/24 09:16

Subject Name

RFI

Subject I.D.

Cpl. B. Owens 1423

Operator Name, I.D.

Permit 230048

Location

504 SE Blue Parkway

Leas Summit No



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BENJAMIN A. OWENS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/27/2023

NUMBER 230048

EXPIRES 3/27/2025

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Douglas J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator OWENS, BENJAMIN
Permit No 230048
Date Issued 3/27/2023 **Date Expires** 3/27/2025

