



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 10:56 am, Jan 13, 2025

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111755	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 12/18/2024
------------------------------------	--	---

LOCATION OF INSTRUMENT (STREET AND CITY) Troop G Headquarters, Willow Springs	TIME OF INSPECTION 11:10 AM
---	---------------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **23⁰c**
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER Guth Laboratories Inc LOT # 23390 EXP. DATE 10/17/2025	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2473 SIM. NIST EXP DATE 02/01/2025	

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .099	TEST 2 → .099	TEST 3 → .099
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

RECALIBRATED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Thomas Meyer
TYPE II PERMIT NUMBER/EXPIRATION DATE 240173 08/16/2026	TELEPHONE NUMBER (417-469-3121)

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00691

Temp Date Time 210L
s/

Air Blank:
12/18/24 11:11 .000
Calibration:
23 12/18/24 11:11 .100

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 240173
Location

Trap G HQ

Willow Springs

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00692

Temp Date Time 210L
s/

Air Blank:
12/18/24 11:12 .000
Calibration Check:
23 12/18/24 11:12 .099

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 240173
Location

Trap G HQ

Willow Springs

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00693

Temp Date Time 210L
s/

Air Blank:
12/18/24 11:13 .000
Calibration Check:
23 12/18/24 11:13 .099

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 240173
Location

Trap G HQ

Willow Springs

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00694

Temp Date Time 210L
s/

Air Blank:
12/18/24 11:17 .000
Calibration Check:
24 12/18/24 11:17 .099

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 240173
Location

Trap G HQ

Willow Springs

AS IV Serial no: 111755
Version no: 532B

TEST RECORD 00695

Temp Date Time 2

VOID: RFI
12 12/18/24 11:20

Subject Name

Subject I.D.

Operator Name, I.D.

T. MEYER 240173
Location

Trap G HQ

Willow Springs



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
THOMAS W. MEYER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/16/2024

NUMBER 240173

EXPIRES 8/16/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula F. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MEYER, THOMAS
Permit No 240173
Date Issued 8/16/2024 **Date Expires** 8/16/2026

