



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 8:24 am, Nov 08, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|   |  |                                  |
|---|--|----------------------------------|
| ALCO SENSOR IV SN<br>111767   | NAME OF AGENCY<br>Saint Joseph Police Department | DATE OF INSPECTION<br>11/07/2024 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>501 Faraon Street, Saint Joseph MO. 64501 |  | TIME OF INSPECTION<br>11:01 am   |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG304601 EXP. DATE 02/15/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

|              |              |              |
|--------------|--------------|--------------|
| TEST 1  .101 | TEST 2  .104 | TEST 3  .105 |
|--------------|--------------|--------------|

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |         |           |           |           |            |
|----------|---------|-----------|-----------|-----------|------------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | (OVER .19) |
|----------|---------|-----------|-----------|-----------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

|  |                                    |
|--|------------------------------------|
| SIGNATURE<br>  | PRINT NAME<br>John L. Foster       |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>230163 Exp-08/07/2025 | TELEPHONE NUMBER<br>(816) 596-8206 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

# SAINT JOSEPH POLICE DEPARTMENT MONTHLY MAINTENANCE REPORT ASIV- W/PRINTER

|  |   |  |   |
|--|---|--|---|
| <p>AS IV Serial no: 111767<br/>Version no: 532B</p> <p>TEST RECORD 00677</p> <p>Temp Date Time 210L<br/>g/</p> <p>Air Blank:<br/>11/07/24 11:01 .000</p> <p>Calibration Check:<br/>23 11/07/24 11:01 .101</p> <p>Subject Name<br/><i>Monthly Test</i></p> <p>Subject I.D.<br/><i>Foster Mark 28977</i></p> <p>Operator Name, I.D.<br/><i>501 Farva Stant</i></p> <p>Location<br/><i>501 + Joseph Mall 4509</i></p> <p style="text-align: center;"><i>LEC</i></p> | <p>AS IV Serial no: 111767<br/>Version no: 532B</p> <p>TEST RECORD 00678</p> <p>Temp Date Time 210L<br/>g/</p> <p>Air Blank:<br/>11/07/24 11:03 .000</p> <p>Calibration Check:<br/>24 11/07/24 11:03 .104</p> <p>Subject Name<br/><i>Monthly Test</i></p> <p>Subject I.D.<br/><i>Foster Mark 28977</i></p> <p>Operator Name, I.D.<br/><i>501 Farva Stant</i></p> <p>Location<br/><i>LEC</i></p> | <p>AS IV Serial no: 111767<br/>Version no: 532B</p> <p>TEST RECORD 00679</p> <p>Temp Date Time 210L<br/>g/</p> <p>Air Blank:<br/>11/07/24 11:04 .000</p> <p>Calibration Check:<br/>24 11/07/24 11:04 .105</p> <p>Subject Name<br/><i>Monthly Testing</i></p> <p>Subject I.D.<br/><i>Foster Mark 28977</i></p> <p>Operator Name, I.D.<br/><i>501 Farva Stant</i></p> <p>Location<br/><i>LEC</i></p> | <p>AS IV Serial no: 111767<br/>Version no: 532B</p> <p>TEST RECORD 00680</p> <p>Temp Date Time 210L<br/>g/</p> <p>VOID: RFI<br/>12 11/07/24 11:06</p> <p>Subject Name<br/><i>Monthly Test</i></p> <p>Subject I.D.<br/><i>Foster Mark 28977</i></p> <p>Operator Name, I.D.<br/><i>501 Farva Stant</i></p> <p>Location<br/><i>LEC</i></p> |
|--|---|--|---|



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 15-Feb-2023

**Lot #** AG304601 **Model** 108

|                                |                         |   |   |
|--------------------------------|-------------------------|---|---|
| <b>Exp Date</b><br>15-Feb-2025 | <b>Cyl. Type</b><br>108 | <b>Component</b><br>Ethanol<br>Nitrogen | <b>Certified Concentration</b><br>0.100 ± 2% BrAC (272 ppm) |
|--------------------------------|-------------------------|---|---|

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581      | 391.8 ppm     | EB0010603      | 392.5 ppm     |
| EB0010570      | 259.8 ppm     | EB0010559      | 258.9 ppm     |
| EB0010285      | 209.0 ppm     | EB0010562      | 104.2 ppm     |
| EB0010561      | 103.7 ppm     | EB0010579      | 52.94 ppm     |
| EB0010681      | 52.22 ppm     |                |               |

  

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481       | 800.0 ppm     | CC727493       | 390.0 ppm     |
| CC727496       | 253.0 ppm     | CC727498       | 150.0 ppm     |

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:02.16.2023 13:50

Approved for Release:   
 Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JOHN L. FOSTER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/7/2023

NUMBER 230163

EXPIRES 8/7/2025

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Douglas J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

**Operator** FOSTER, JOHN  
**Permit No** 230163  
**Date Issued** 8/7/2023    **Date Expires** 8/7/2025

