

**RECEIVED**  
By Tracy Crews at 2:54 pm, Aug 16, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111760      NAME OF AGENCY JOHNSON COUNTY SHERIFF'S OFFICE      DATE OF INSPECTION 08/16/2024

LOCATION OF INSTRUMENT (STREET AND CITY) 278 SOUTH WEST 871 ROAD CENTERVIEW, MISSOURI 64019      TIME OF INSPECTION 9:08 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION       COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS      LOT # AG310903      EXP. DATE 04/19/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C)      SIM. SN      SIM. NIST EXP DATE

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099      TEST 2 .098      TEST 3 .097

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT. (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS    0    (0-.04)    0    (.05-.09)    1    (.10-.14)    (.15-.19)    0    (OVER .19)    3

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE [Signature]      PRINT NAME PATRICK W. BOYER

TYPE (PERMIT NUMBER/EXPIRATION DATE) 230100 / 05-30-2025      TELEPHONE NUMBER (660) 747-6469

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111760  
Version no: 532B

TEST RECORD 00834

Temp Date Time 210L

Air Blank:  
08/16/24 09:08 .000  
Calibration Check:  
19 08/16/24 09:08 .099

Subject Name  
TEST #1

Subject I.D.

Operator Name, I.D.

Bayer 230600

Location  
JCSO

AS IV Serial no: 111760  
Version no: 532B

TEST RECORD 00835

Temp Date Time 210L

Air Blank:  
08/16/24 09:10 .000  
Calibration Check:  
21 08/16/24 09:10 .098

Subject Name  
TEST #2

Subject I.D.

Operator Name, I.D.

Bayer 230600

Location  
JCSO

AS IV Serial no: 111760  
Version no: 532B

TEST RECORD 00836

Temp Date Time 210L

Air Blank:  
08/16/24 09:11 .000  
Calibration Check:  
21 08/16/24 09:11 .097

Subject Name  
TEST #3

Subject I.D.

Operator Name, I.D.

Bayer 230600

Location  
JCSO

AS IV Serial no: 111760  
Version no: 532B

TEST RECORD 00837

Temp Date Time 210L

VOID: RFI  
12 08/16/24 09:12

Subject Name  
TEST RFI

Subject I.D.

Operator Name, I.D.

Bayer 230600

Location  
JCSO

AS IV Serial no: 111760  
Version no: 532B

TEST RECORD 00838

Temp Date Time 210L

Air Blank:  
08/16/24 09:14 .000  
Subject Test: Auto  
23 08/16/24 09:14 .000


Subject Name  
SELF TEST

Subject I.D.

Operator Name, I.D.

Bayer 230600

Location  
JCSO



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

### Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 20-Apr-2023

**Lot #** AG310903 **Model** 108

<b>Exp Date</b> 19-Apr-2025	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
Reason:Dry gas standard certification of analysis  
Location:Airgas USA LLC (Lab)  
Date:04.20.2023 15:24

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**PATRICK W. BOYER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/30/2023

NUMBER 230100

EXPIRES 5/30/2025

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **BOYER, PATRICK**  
Permit No **230100**  
Date Issued **5/30/2023** Date Expires **5/30/2025**