

RECEIVED

By Tracy Crews at 7:41 am, Nov 18, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN **111780** NAME OF AGENCY **Kansas City Police Department** DATE OF INSPECTION **11/15/2024**

LOCATION OF INSTRUMENT (STREET AND CITY) **9701 MARION PARK DR, KANSAS CITY** TIME OF INSPECTION **1843**

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **INTOXIMETERS** LOT # **AG324501** EXP. DATE **09/02/2025**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 **.098** TEST 2 **.099** TEST 3 **.099**

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS **0** (0-.04) **0** (.05-.09) **0** (.10-.14) **0** (.15-.19) **0** (OVER .19) **0**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE *P.O. Eric Lyles #50045*

PRINT NAME **Eric Lyles**

TYPE II PERMIT NUMBER/EXPIRATION DATE **240189- 08/29/2026**

TELEPHONE NUMBER **816-716-4241**

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00979

Temp Date Time ^{g/} 210L

Air Blank:
11/15/24 18:43 .000
Calibration Check:
21 11/15/24 18:43 .098

Subject Name

Test # 1

Subject I.D.

Operator Name, I.D.

Lyles #240189

Location

9701 Marmon Park Dr

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00980

Temp Date Time ^{g/} 210L

Air Blank:
11/15/24 18:44 .000
Calibration Check:
21 11/15/24 18:44 .099

Subject Name

Test # 2

Subject I.D.

Operator Name, I.D.

Lyles #240189

Location

9701 Marmon Park Dr

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00981

Temp Date Time ^{g/} 210L

Air Blank:
11/15/24 18:46 .000
Calibration Check:
22 11/15/24 18:46 .099

Subject Name

Test # 3

Subject I.D.

Operator Name, I.D.

Lyles #240189

Location

9701 Marmon Park Dr

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00982

Temp Date Time ^{g/} 210L

VOID: RFI
12 11/15/24 18:47

Subject Name

Test # 4 RFI

Subject I.D.

Operator Name, I.D.

Lyles #240189

Location

9701 Marmon Park Dr



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

ERIC LYLES

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/29/2024

NUMBER 240189

EXPIRES 8/29/2026

MS 59-0771 (6-01)

Mike Morrison
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Redden
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an individual breath alcohol analyzer for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: **LYLES, ERIC**
Permit No: **240189**
Date Issued: **8/29/2024** Date Expires: **8/29/2026**



Airgas USA LLC (LAB)
2600 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Test Date: 25-Jun-2024

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Lot # AG417401 Model 108

Exp Date 22-Jun-2026 Cyl. Type 108 Certified Concentration 0.100 ± 2% BrAC (27.2 ppm)
Component Ethanol Nitrogen

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	CRM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: NDIR

Quality Approved by National Control Laboratory
London, Ontario, Canada
L6M 3V3
L6M 3V3 (USA, LLC (LAB))
Date: 03/25/2024 13:25

Yusef Woods

Approved for Release: _____
Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.00
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07