

RECEIVED

By Tracy Crews at 8:45 am, Jul 08, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 119726	NAME OF AGENCY Chesterfield Police Department	DATE OF INSPECTION 07/04/2024
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 690 Chesterfield Pkwy W, Chesterfield	TIME OF INSPECTION 2:01 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG223501 EXP. DATE 08/23/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.080

TEST 2 0.080

TEST 3 0.080

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within Missouri Department of Health Specifications.
Test #00106 was RFI test that did not print due to low power on printer. Test #00107 was second RFI test.

INSPECTING OFFICER

SIGNATURE

PRINT NAME
P.O. Rupp #264

TYPE | PERMIT NUMBER | EXPIRATION DATE
230076 / 04/14/2025

TELEPHONE NUMBER
(636) 537-3000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119726
Version no: 532C

TEST RECORD 00103
Temp Date Time 9/
Air Blank: 07/04/24 14:01 .000
Calibration Check: 23 07/04/24 14:01 .080

Subject Name
Test #1
Subject I.D.
G. Rupp # 264
Operator Name, I.D.
230076
Location
Chestnut Hill PD

AS IV Serial no: 119726
Version no: 532C

TEST RECORD 00104
Temp Date Time 9/
Air Blank: 07/04/24 14:03 .000
Calibration Check: 23 07/04/24 14:03 .080

Subject Name
Test #2
Subject I.D.
G. Rupp # 264
Operator Name, I.D.
230076
Location
Chestnut Hill PD

AS IV Serial no: 119726
Version no: 532C

TEST RECORD 00105
Temp Date Time 9/
Air Blank: 07/04/24 14:04 .000
Calibration Check: 23 07/04/24 14:04 .080

Subject Name
Test #3
Subject I.D.
G. Rupp # 264
Operator Name, I.D.
230076
Location
Chestnut Hill PD

AS IV Serial no: 119726
Version no: 532C

TEST RECORD 00107
Temp Date Time 9/
VOID: RFI
12 07/04/24 14:06

Subject Name
RFI Test
Subject I.D.
G. Rupp # 264
Operator Name, I.D.
230076
Location
Chestnut Hill PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
GREGORY RUPP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/14/2023

NUMBER 230076

EXPIRES 4/14/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RUPP, GREGORY
Permit No 230076
Date Issued 4/14/2023 **Date Expires** 4/14/2025

