



RECEIVED

By Tracy Crews at 7:35 am, Jul 30, 2024 #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 119736	NAME OF AGENCY LAWRENCE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 07/29/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 240 N. MAIN ST - MT. VERNON, MO 65712		TIME OF INSPECTION 8:59 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG333203 EXP. DATE 11/28/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .096	TEST 2 → .096	TEST 3 → .095
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME RYAN DEVOST
TYPE II PERMIT NUMBER/EXPIRATION DATE 230066 - 04/07/2025	TELEPHONE NUMBER (417) 466-2131

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119730
Version no: 532C

AS IV Serial no: 119730
Version no: 532C

AS IV Serial no: 119730
Version no: 532C

TEST RECORD 00200

TEST RECORD 00201

TEST RECORD 00202

Temp Date Time 2167

Temp Date Time 2167

Temp Date Time 2167

Air Blank:
07/29/24 20:59 .000
Calibration Check:
28 07/29/24 20:59 .000

Air Blank:
07/29/24 21:01 .000
Calibration Check:
28 07/29/24 21:01 .000

Air Blank:
07/29/24 21:02 .000
Calibration Check:
29 07/29/24 21:02 .000

Subject Name

Subject Name

Subject Name

TEST #1

TEST #2

TEST #3

Subject I.D.

Subject I.D.

Subject I.D.

Operator Name, I.D.



Operator Name, I.D.

Operator Name, I.D.

Location

Location

Location



AS IV Serial no: 119730
Version no: 532C

AS IV Serial no: 119730
Version no: 532C

TEST RECORD 00203

TEST RECORD 00204

Temp Date Time 2173

Temp Date Time 2173

Air Blank:
07/29/24 21:04 .000
Calibration Check:
29 07/29/24 21:04 .000

VOID REF:
12 07/29/24 21:05

Subject Name

Subject Name

SOBER Sample

REF TEST

Subject I.D.



Subject I.D.

Operator Name, I.D.

Operator Name, I.D.

Location

Location





Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 29-Nov-2023

Lot # AG333203 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
28-Nov-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 11.30.2023 17:29

Approved for Release: _____

Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RYAN DEVOST

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s)

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit limited under the provisions of sections 577.020 through 577.041, RSMo and 500.111 through 500.119 RSMo.

DATE 11/7/2023

Michael M...
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230066

David M. Richardson

EXPIRES 11/7/2025

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
L.A. 01 (10-18)

Mo 500.071 (b-1b)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The permit holder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath (or) of expired air in Missouri.

Operator DEVOST, RYAN
Permit No 200066
Date Issued 11/7/2023 Date Expires 11/7/2025

