



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12820	NAME OF AGENCY WASHINGTON POLICE DEPT	DATE OF INSPECTION 11/05/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 301 Jefferson St. Washington		TIME OF INSPECTION 03:27 CST

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST
<input checked="" type="checkbox"/> ETH CHECK	

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER INTOXIMETER	LOT# AG305102 EXP. DATE 02/20/2025
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

<input checked="" type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.100 g/210L	TEST 2 0.100 g/210L	TEST 3 0.099 g/210L
---------------------	---------------------	---------------------

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS	0	0-.04	20	.05-.09	0	.10-.14	2	.15-.19	1	OVER .19	0
----------	---	-------	----	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE Sgt. D. #285	PRINT FULL NAME TOLLISON, DOUGLAS
TYPE II PERMIT NUMBER 230055	EXPIRATION DATE 03/27/2025
	TELEPHONE NUMBER (636) 390-1050

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, Missouri Department of Health and Senior Services,
 by mail, fax, or e-mail

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 20-Feb-2023

Lot # AG305102 **Model** 108

Exp Date 20-Feb-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	-----------------------------------------	-------------------------------------------------------------

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010550	258.9 ppm
EB0010286	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727488	253.0 ppm	CC727488	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)
Date: 02-20-2023 11:27:37

Approved for Release:



Rod Margala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

STATE OF MISSOURI)
)
COUNTY OF FRANKLIN) SS

AFFIDAVIT

Before me, the undersigned authority, personally appeared, Douglas A. Tollison who, being by me duly sworn, deposed as follows:

My name is Douglas A. Tollison I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of maintenance records of the INTOX EC/IR II, SERIAL #12820. Attached hereto are 3 pages of records from the Washington Police Department for the month of November, 202024. These 3 pages of records are kept by Washington Police Department in the regular course of business, and it was the regular course of business of Washington Police Department for an employee or representative of Washington Police Department with knowledge of the act, event, condition, opinion or diagnosis recorded to make the record or to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion or diagnosis. The records attached hereto are the original or exact duplicate of the original.

SGT.  #1285


In witness whereof I have hereunto subscribed my name and affixed my official seal this 6 day of November, 2024.

STEPHANIE L. HELLEBUSCH
NOTARY PUBLIC - NOTARY SEAL
STATE OF MISSOURI
MY COMMISSION EXPIRES MAY 11, 2025
FRANKLIN COUNTY
COMMISSION #11134383

Stephanie L. Hellebusch
Notary Public



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

DOUGLAS A. TOLLISON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/27/2023

NUMBER 230055

EXPIRES 3/27/2025

M. L. Morrison

COMMISSIONER, STATE DEPARTMENT OF HEALTH AND SENIOR SERVICES

Paula J. Nicholson

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

Operator: TOLLISON DOUGLAS
Permit No: 230055
Date issued 3/27/2023 Date Expires 3/27/2025

