

RECEIVED

By Tracy Crews at 7:28 am, Sep 25, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|--|----------------------------------|
| INTOX DMT SN 500020 | NAME OF AGENCY Peculiar Police Department | DATE OF INSPECTION 09/24/2024 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 224 N. Main Peculiar, MO 64078 | | TIME OF INSPECTION 23:37:55 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

DIAGNOSTIC RECORD

| | |
|--|---|
| DATE AND TIME <u>09/24/2024 23:37:57</u> | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTER 1 |
| <input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u> | <input checked="" type="checkbox"/> FILTER 2 |
| <input checked="" type="checkbox"/> BREATH TUBE <u>45.5°C</u> | <input checked="" type="checkbox"/> FILTER 3 |
| <input checked="" type="checkbox"/> PUMP | <input checked="" type="checkbox"/> INTERNAL STANDARD |

BREATH ANALYZER ACCURACY STANDARDS

SIMULATOR STANDARD COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 SIM. SN SD2260 SIM. NIST EXP DATE 01/19/2025

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.097 TEST 2: 0.098 TEST 3: 0.099

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

| | | | | | |
|-------------|----------|------------|------------|------------|-------------|
| REFUSALS: 0 | 0-.04: 0 | .05-.09: 0 | .10-.14: 1 | .15-.19: 0 | OVER .19: 0 |
|-------------|----------|------------|------------|------------|-------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

| | | |
|---------------------------------|--------------------------------------|----------------------------------|
| SIGNATURE | PRINT FULL NAME ALEXANDER GOODMAN | |
| TYPE II PERMIT NUMBER 230246 | EXPIRATION DATE 10/31/2025 | TELEPHONE NUMBER 816-779-5102 |

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services by mail, fax, or email



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

ALEXANDER B. GOODMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/31/2023

NUMBER 230246

EXPIRES 10/31/2025

Mile Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-19)

MO-493 (04/2016)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **GOODMAN, ALEXANDER**
Permit No **230246**
Date Issued **10/31/2023** Date Expires **10/31/2025**

