

RECEIVED

By Tracy Crews at 8:05 am, Dec 04, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|--|--|---|
| INTOX DMT SN 500058 | NAME OF AGENCY Missouri State Highway Patrol | DATE OF INSPECTION 12/03/2024 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 106 N. Folger St., Carrollton, MO 64633 | | TIME OF INSPECTION 18:32:59 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

| | |
|---|--|
| <input checked="" type="checkbox"/> DIAGNOSTIC RECORD | |
| DATE AND TIME <u>12/03/2024 18:33:02</u> | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTER 1 |
| <input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u> | <input checked="" type="checkbox"/> FILTER 2 |
| <input checked="" type="checkbox"/> BREATH TUBE <u>48.1°C</u> | <input checked="" type="checkbox"/> FILTER 3 |
| <input checked="" type="checkbox"/> PUMP | <input checked="" type="checkbox"/> INTERNAL STANDARD |

| | |
|--|---|
| BREATH ANALYZER ACCURACY STANDARDS | |
| <input type="checkbox"/> SIMULATOR STANDARD | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u> | LOT # <u>AG234103</u> EXP. DATE <u>12/07/2024</u> |
| <input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) | SIM. SN _____ SIM. NIST EXP DATE _____ |

| |
|--|
| <input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used. |
| <input checked="" type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE |
| <input type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE |
| <input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE |

| | | |
|----------------------|----------------------|----------------------|
| TEST 1: 0.098 | TEST 2: 0.098 | TEST 3: 0.098 |
|----------------------|----------------------|----------------------|

| |
|--|
| <input checked="" type="checkbox"/> PERFORM R.F.I. TEST |
|--|

| | | | | | |
|---|-----------------|-------------------|-------------------|-------------------|--------------------|
| INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: | | | | | |
| REFUSALS: 0 | 0-.04: 0 | .05-.09: 0 | .10-.14: 0 | .15-.19: 0 | OVER .19: 0 |

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

| | | |
|--|---|---|
| SIGNATURE | PRINT FULL NAME ISAAC L KIMBALL | |
| TYPE II PERMIT NUMBER 230322 | EXPIRATION DATE 12/21/2025 | TELEPHONE NUMBER 816-622-0800 |

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services
by mail, fax, or email



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Dec-2022

Lot # AG234103 **Model** 108

| | | | |
|-----------------|------------------|---------------------|--------------------------------|
| Exp Date | Cyl. Type | Component | Certified Concentration |
| 7-Dec-2024 | 108 | Ethanol Nitrogen | 0.100 ± 2% BrAC (260 ppm) |

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010562 | 104.2 ppm |
| EB0010561 | 103.7 ppm | EB0010579 | 52.94 ppm |
| EB0010681 | 52.22 ppm | | |

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481 | 800.0 ppm | CC727493 | 390.0 ppm |
| CC727496 | 253.0 ppm | CC727498 | 150.0 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 12.09.2022 17:20

Approved for Release: 
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ISAAC KIMBALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

NUMBER 230322

EXPIRES 12/21/2025

Mike Maszura

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator KIMBALL, ISAAC
Permit No 230322
Date Issued 12/21/2023 **Date Expires** 12/21/2025

