

RECEIVED

By Tracy Crews at 12:23 pm, Nov 06, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|------------------------|-------------------------------------------------|----------------------------------|
| INTOX DMT SN 500060 | NAME OF AGENCY Missouri State Highway Patrol | DATE OF INSPECTION 11/01/2024 |
|------------------------|-------------------------------------------------|----------------------------------|

| | |
|-------------------------------------------------------------------------------------|--------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) Taney County Sheriff's Office - Forsyth | TIME OF INSPECTION 09:08:45 |
|-------------------------------------------------------------------------------------|--------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

| | |
|------------------------------------------------------------------|-------------------------------------------------------|
| <input checked="" type="checkbox"/> DIAGNOSTIC RECORD | |
| DATE AND TIME <u>11/01/2024 09:08:47</u> | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTER 1 |
| <input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.7°C</u> | <input checked="" type="checkbox"/> FILTER 2 |
| <input checked="" type="checkbox"/> BREATH TUBE <u>39.9°C</u> | <input checked="" type="checkbox"/> FILTER 3 |
| <input checked="" type="checkbox"/> PUMP | <input checked="" type="checkbox"/> INTERNAL STANDARD |

| | |
|---------------------------------------------|--------------------------------------------------------------------|
| BREATH ANALYZER ACCURACY STANDARDS | |
| <input type="checkbox"/> SIMULATOR STANDARD | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |

| | | |
|---------------------------------------------------------------------------|-----------------------|-----------------------------|
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u> | LOT # <u>AG335303</u> | EXP. DATE <u>12/19/2025</u> |
| <input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) | SIM. SN | SIM. NIST EXP DATE |

| |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used. |
| <input checked="" type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE |
| <input type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE |
| <input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE |

| | | |
|---------------|---------------|---------------|
| TEST 1: 0.097 | TEST 2: 0.097 | TEST 3: 0.097 |
|---------------|---------------|---------------|

| |
|---------------------------------------------------------|
| <input checked="" type="checkbox"/> PERFORM R.F.I. TEST |
|---------------------------------------------------------|

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

| | | | | | |
|-------------|----------|------------|------------|------------|-------------|
| REFUSALS: 1 | 0-.04: 1 | .05-.09: 0 | .10-.14: 3 | .15-.19: 5 | OVER .19: 1 |
|-------------|----------|------------|------------|------------|-------------|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

| |
|--|
| |
| |
| |
| |
| |
| |

INSPECTING OFFICER

| | |
|------------------------------|-----------------------------|
| SIGNATURE <i>B A Hall</i> | PRINT FULL NAME B A HALL |
|------------------------------|-----------------------------|

| | | |
|---------------------------------|-------------------------------|----------------------------------|
| TYPE II PERMIT NUMBER 230011 | EXPIRATION DATE 01/24/2025 | TELEPHONE NUMBER 417-895-6868 |
|---------------------------------|-------------------------------|----------------------------------|

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services by mail, fax, or email



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRANDON A. HALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/24/2023

NUMBER 230011

EXPIRES 1/24/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HALL, BRANDON
Permit No 230011
Date Issued 1/24/2023 **Date Expires** 1/24/2025

