

**RECEIVED**

By Tracy Crews at 9:19 am, Jul 09, 2024



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**INTOX DMT MAINTENANCE REPORT**

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500068	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/01/2024
------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Christian Co S.O. - 110 W. Elm St., Ozark	TIME OF INSPECTION 08:19:32
---	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

 **DIAGNOSTIC RECORD**DATE AND TIME 07/01/2024 08:19:36 DETECTOR PROGRAM FILTER 1 SAMPLE CHAMBER 48.7°C FILTER 2 BREATH TUBE 48.1°C FILTER 3 PUMP INTERNAL STANDARD**BREATH ANALYZER ACCURACY STANDARDS** SIMULATOR STANDARD COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER INTOXIMETERS LOT # AG335303 EXP. DATE 12/19/2025 SIMULATOR TEMP (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.101

TEST 2: 0.101

TEST 3: 0.101

 **PERFORM R.F.I. TEST****INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS: 0

0-.04: 5

.05-.09: 0

.10-.14: 0

.15-.19: 0

OVER .19: 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

July Maintenance

**INSPECTING OFFICER**

SIGNATURE

PRINT FULL NAME

JACKSON D HOOD

TYPE II PERMIT NUMBER

230167

EXPIRATION DATE

08/07/2025

TELEPHONE NUMBER

417-895-6868

RETURN COMPLETED REPORT TO THE

Breath Alcohol Program, Missouri Department of Health and Senior Services  
by mail, fax, or email



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 19-Dec-2023

**Lot #** AG335303 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
19-Dec-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 12.21.2023 20:20

**Approved for Release:** \_\_\_\_\_  
 Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II  
 JACKSON HOOD**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/7/2023

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230167

*Douglas J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/7/2025

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator HOOD, JACKSON  
 Permit No 230167  
 Date Issued 8/7/2023 Date Expires 8/7/2025

