


By Tracy Crews at 8:09 am, May 20, 2024

RECEIVED



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
INTOXICANT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular or only preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 30 days to the Breath Alcohol Program, DHSS.

INTOXICANT SN 500260	NAME OF AGENCY KENNESAW POLICE DEPARTMENT	DATE OF INSPECTION 05/14/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 200 Cedar St. Kennesaw, Mo 63857		TIME OF INSPECTION 14:47:09

CHECKLIST: Place a mark in the box by the instrument if found to be satisfactory or is operating within established limits. (Write in observed values where determining.) Unmarked items are to be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC RECORD	
DATE AND TIME: 05/14/2024 14:47:09	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER: 18.9°C	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE: 16.6°C	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input type="checkbox"/> SIMULATOR STANDARD	

<input checked="" type="checkbox"/> STANDARD SUPPLY FOR INTOXIMETER	LOT # AG220102	EXP. DATE 07/20/2024
<input type="checkbox"/> SIMULATOR TEMP (41°C ± 0.2°C)	M. SN	SIM. NIST EXP DATE

<input checked="" type="checkbox"/> CALIBRATION CHECK - ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT	Run three tests using a standard. All three must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.
<input checked="" type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.073% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.033% AND 0.042% INCLUSIVE	

TEST 1: 0.100	TEST 2: 0.099	TEST 3: 0.100
---------------	---------------	---------------

<input checked="" type="checkbox"/> PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 3	0-9: 2	10-14: 1	15-19: 0	OVER 19: 1
-------------	--------	----------	----------	------------

LIST ANY NEW PARTS AND DEVICES, AND ALTERATION OR REPAIRS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME ALAN B CAMPBELL
-----------	------------------------------------

TYPE II PERMIT NUMBER 230091	EXPIRATION DATE 05/22/2025	TELEPHONE NUMBER 573-888-4622
---------------------------------	-------------------------------	----------------------------------

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services
mail, fax, or email



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

ALAN CAMPBELL

is hereby authorized to instruct, supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath alcohol instrument(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo. and 306.111 through 306.119 RSMo.

DATE 5/22/2021

NUMBER 230091

EXPIRES 5/22/2025

MO 680-0771 (6-10)

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

Permit holder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air

CAMPBELL, ALAN
230091
Date Issued 5/22/2021 Date Expires 5/22/2025