



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

RECEIVED  
 BY: [Name] DATE: [Date]

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005847	LOCATION OF INSTRUMENT IPD BATVAN STATION 1	DATE OF INSPECTION 11/14/2024	TIME OF INSPECTION 16:25
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG306503	STANDARD EXPIRATION DATE 03/06/2025
Air Blank	0.000	16:27	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.098	16:27	STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS, INC	
Air Blank	0.000	16:28	CALIBRATION CHECK RESULT 1 0.098		
Cal Check	0.098	16:28	CALIBRATION CHECK RESULT 2 0.098		
Air Blank	0.000	16:29	CALIBRATION CHECK RESULT 3 0.098		
Cal Check	0.098	16:29	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.0%	SPREAD (MUST BE .005 OR LESS) 0.000	
Air Blank	0.000	16:29	<h1>Pass</h1>		

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	16:30
EEPROM Checksum Test	Pass		Subject Test	RFI*	16:30
Real Time Clock Test	Pass		Air Blank	0.000	16:31
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass		<h1>Pass</h1>		
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT OPERATED WITHIN STATE STANDARDS

INSPECTING OFFICER		
SIGNATURE	PRINT NAME RON BALTZER	
TYPE II PERMIT NUMBER 230132	EXPIRATION DATE 07/06/2025	TELEPHONE NUMBER 816-325-7300



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Test Date: 7-Mar-2023

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Lot # AG306503 Model 108

Exp Date	Cyl. Type	Component	Certified Concentration
6-Mar-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:


RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 03.08.2023 20:47

Approved for Release:   
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**RON BALTZER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DATE 7/6/2023

NUMBER 230132

EXPIRES 7/6/2025

*David J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BALTZER, RON  
 Permit No 230132  
 Date Issued 7/6/2023 Date Expires 7/6/2025

