



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005854	LOCATION OF INSTRUMENT CLAY COUNTY, MO DET.	DATE OF INSPECTION 07/03/2024	TIME OF INSPECTION 08:32
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	08:34	DRY	AG407801	03/18/2026
Cal Check	0.105	08:35	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	08:35	N/A	N/A	N/A
Cal Check	0.105	08:36	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	08:36	0.100	INTOXIMETERS	
Cal Check	0.105	08:36	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	08:36	0.105		
Cal Check	0.105	08:36	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	08:37	0.105		
<b>Pass</b>			CALIBRATION CHECK RESULT 3		
			0.105		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			5.0%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	08:37
RAM Test	Pass		Subject Test	RFI*	08:38
EEPROM Checksum Test	Pass		Air Blank	0.000	08:38
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		<b>Pass</b>		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	0	0	0	0	0	2

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

PASS. NEW DRY STANDARD

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME BENJAMIN LAUGHLIN	
TYPE II PERMIT NUMBER 230106	EXPIRATION DATE 05/30/2025	TELEPHONE NUMBER 816.439.4701	