



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
07/18/2024 09:47:42 AM

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

| | | | |
|---------------------------------------|------------------------------------------------|----------------------------------|-----------------------------|
| INSTRUMENT SERIAL NUMBER 80-006195 | LOCATION OF INSTRUMENT CARTHAGE POLICE DEPT | DATE OF INSPECTION 07/20/2024 | TIME OF INSPECTION 07:14 |
|---------------------------------------|------------------------------------------------|----------------------------------|-----------------------------|

| CALIBRATION CHECK RESULTS | | | CALIBRATION CHECK SUMMARY | | |
|---------------------------|--------|-------|-----------------------------------------------|----------------------------------------|----------------------------------------|
| Test | g/210L | Time | STANDARD TYPE DRY | STANDARD LOT # AG222301 | STANDARD EXPIRATION DATE 08/11/2024 |
| Air Blank | 0.000 | 07:15 | SIM TEMPERATURE N/A | SIM SERIAL NUMBER N/A | SIM CERTIFICATE EXPIRATION N/A |
| Cal Check | 0.100 | 07:15 | STANDARD VALUE 0.100 | STANDARD SUPPLIER INTOXIMETERS | |
| Air Blank | 0.000 | 07:16 | CALIBRATION CHECK RESULT 1 0.100 | | |
| Cal Check | 0.100 | 07:16 | CALIBRATION CHECK RESULT 2 0.100 | | |
| Air Blank | 0.000 | 07:17 | CALIBRATION CHECK RESULT 3 0.101 | | |
| Cal Check | 0.101 | 07:17 | MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.0% | SPREAD (MUST BE .005 OR LESS) 0.001 | |
| Air Blank | 0.000 | 07:17 | <h1>Pass</h1> | | |

| DIAGNOSTIC TEST RESULTS | | | RFI TEST RESULTS | | |
|-----------------------------|------|--|------------------|--------|-------|
| Voltage/Current Test | Pass | | Test | g/210L | Time |
| RAM Test | Pass | | Air Blank | 0.000 | 07:18 |
| EEPROM Checksum Test | Pass | | Subject Test | RFI* | 07:18 |
| Real Time Clock Test | Pass | | Air Blank | 0.000 | 07:19 |
| DSP Test | Pass | | *RFI Detect | | |
| Analytical Stability Test | Pass | | <h1>Pass</h1> | | |
| Modem Test | Pass | | | | |
| Temperature Regulation Test | Pass | | | | |

| NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT | | | | | | | |
|-----------------------------------------------------------------------------------------|---------|---------|---------|---------|----------|--|--|
| REFUSALS | .00-.04 | .05-.09 | .10-.14 | .15-.19 | OVER .19 | | |
| 0 | 0 | 0 | 1 | 1 | 1 | | |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

JUNE 2024 MAINTENANCE

| INSPECTING OFFICER | | | |
|---------------------------------|-------------------------------|--------------------------------|--|
| SIGNATURE | | PRINT NAME BRADY LUCE | |
| TYPE II PERMIT NUMBER 240106 | EXPIRATION DATE 05/14/2026 | TELEPHONE NUMBER 4172377200 | |



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
 TYPE II
 BRADY LUCE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/14/2024

NUMBER 240106

EXPIRES 5/14/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LUCE, BRADY
 Permit No 240106
 Date Issued 5/14/2024 Date Expires 5/14/2026

