



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007513	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 07/31/2024	TIME OF INSPECTION 23:14
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	23:16	DRY	14323080A4	06/05/2025
Cal Check	0.079	23:16	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	23:17	N/A	N/A	N/A
Cal Check	0.078	23:17	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	23:18	0.080	CMI INC	
Cal Check	0.079	23:18	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	23:18	0.079		
Cal Check	0.079	23:18	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	23:18	0.078		
Pass			CALIBRATION CHECK RESULT 3		
			2.5%		

RECEIVED
By Tracy Crews at 8:40 am, Aug 15, 2024

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	23:19
RAM Test	Pass		Subject Test	RFI*	23:19
EEPROM Checksum Test	Pass		Air Blank	0.000	23:20
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	1	0	0	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
TESTED AND CERTIFIED

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME WADE ROBINSON

TELEPHONE NUMBER 8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**
WADE ROBINSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/23/2023

NUMBER 230228

EXPIRES 10/23/2025

MO 880-0771 (8-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Robinson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LS4-PR-10

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidentiary breath alcohol instrument for the determination of the alcoholic content of breath from an operator.

Operator: ROBINSON, WADE
Permit No. 230228
Date Issued: 10/23/2023 Date Expires: 10/23/2025

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-245-7634 • www.ilmoproducts.com



Certificate of Analysis

Certificate ID: 15762
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14323080A4
Expiration: 6/5/2025

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	288 ppm	+/- 0.002 BAC(72a1)	NDR
Nitrogen	Balance	[5.2 ppm]	

Distributed by:
CPI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 261.0 µmol/lol
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
Blair

06-C1-2023
Issuance Date



ISO/IEC 17025:2017 Accredited Laboratory

The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST and fully compliant with the requirements of the International Vocabulary of Basic and General Terms in Metrology (VIM) and the International System of Units (SI). The accuracy of the results is dependent on the quality of the sample and the quality of the instrument used for the analysis. The instrument used for the analysis is identified in the certificate. The instrument used for the analysis is identified in the certificate. The instrument used for the analysis is identified in the certificate.