



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 7:34 am, Jan 24, 2025

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 035485	PRINTER SN 096.3580.926	DATE OF INSPECTION 01/23/2025
LOCATION OF INSTRUMENT (STREET AND CITY) SCCPD / 101 Sheriff Dierker Court, O'Fallon, Missouri 63366		TIME OF INSPECTION 7:55 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG407801 EXP. DATE 03/18/2026

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .098	TEST 2  .098	TEST 3  .097
--------------	--------------	--------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE #671	PRINT NAME Dane Kaltenbach
TYPE II PERMIT NUMBER/EXPIRATION DATE 230104/ 05/30/2025	TELEPHONE NUMBER (636) 949-3000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS 10 Serial no. 605485  
Version no. 5070

TEST RECORD 00465

Temp Date Time 210L 5/

Air Blinds

01/22/25 19:54 .000

Calibration Check

19 01/22/25 10:14 .099

Subsect 100

TEST

Subject #1

Operator Name  
Kaltenbach / 230104

Location  
SCCPD

AS 10 Serial no. 605485  
Version no. 5070

TEST RECORD 00466

Temp Date Time 210L 5/

Air Blinds

01/22/25 19:54 .000

Calibration Check

19 01/22/25 10:14 .099

Subsect 100

TEST

Subject #2

Operator Name  
Kaltenbach / 230104

Location  
SCCPD

AS 10 Serial no. 605485  
Version no. 5070

TEST RECORD 00472

Temp Date Time 210L 5/

Air Blinds

01/24/25 09:00 .000

Calibration Check

21 01/23/25 01:00 .097

Subsect 100

TEST

Subject #3

Operator Name  
Kaltenbach / 230104

Location  
SCCPD

AS 10 Serial no. 605485  
Version no. 5070

TEST RECORD 00471

Temp Date Time 210L 5/

VOIMS PPT

12 01/23/25 01:00

Subsect 100

TEST

R.F.I.

Operator Name  
Kaltenbach / 230104

Location  
SCCPD

TEST RECORD 00472  
Temp Date Time 210L 5/

Air Blinds

01/24/25 09:00 .000

Calibration Check

21 01/23/25 01:00 .097

Subsect 100

TEST

Subject #3

Operator Name  
Kaltenbach / 230104

Location  
SCCPD

AS 10 Serial no. 605485  
Version no. 5070

TEST RECORD 00471

Temp Date Time 210L 5/

VOIMS PPT

12 01/23/25 01:00

Subsect 100

TEST

R.F.I.

Operator Name  
Kaltenbach / 230104

Location  
SCCPD



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 19-Mar-2024

**Lot #** AG407801 **Model** 108

<b>Exp Date</b> 18-Mar-2026	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	-----------------------------------------	-------------------------------------------------------------

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:03.22.2024 07:50

**Approved for Release:** \_\_\_\_\_  
 Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**DANE KALTENBACH**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **5/30/2023**

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER **230104**

*David J. Nicholson*

EXPIRES **5/30/2025**

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580 2771 (6-10)

LAB 4 (R6 10)

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** KALTENBACH, DANE  
**Permit No** 230104  
**Date Issued** 5/30/2023 **Date Expires** 5/30/2025

