

AS IV Serial no: 087976
Version no: 532B

TEST RECORD 00572

Temp Date Time 210L

Air Blank: 03/05/25 12:05 .000
Calibration Check: 20 03/05/25 12:05 .101

Subject Name TEST 2
Subject I.D.

Operator Name, I.D. DW HENLEY # 727
Location 3131 E KEARNEY

SPRINGFIELD, MO
198 # 727

AS IV Serial no: 087976
Version no: 532B

TEST RECORD 00573

Temp Date Time 210L

Air Blank: 03/05/25 12:07 .000
Calibration Check: 20 03/05/25 12:07 .101

Subject Name TEST 2
Subject I.D.

Operator Name, I.D. DW HENLEY # 727
Location 3131 E KEARNEY

SPRINGFIELD, MO
198 # 727

AS IV Serial no: 087976
Version no: 532B

TEST RECORD 00574

Temp Date Time 210L

Air Blank: 03/05/25 12:09 .000
Calibration Check: 21 03/05/25 12:09 .100

Subject Name TEST 3
Subject I.D.

Operator Name, I.D. DW HENLEY # 727
Location 3131 E KEARNEY

SPRINGFIELD, MO
198 # 727

AS IV Serial no: 087976
Version no: 532B

TEST RECORD 00575

Temp Date Time 210L

VOID: RFI
12 03/05/25 12:10

Subject Name RFI
Subject I.D.

Operator Name, I.D. DW HENLEY # 727
Location 3131 E KEARNEY

SPRINGFIELD, MO
198 # 727



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 24310 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 28, 2024, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is August 27, 2026 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DAVID W. HENLEY, JR.

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/28/2024

NUMBER 240144

EXPIRES 6/28/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HENLEY, JR., DAVID
 Permit No 240144
 Date Issued 6/28/2024 Date Expires 6/28/2026

