

RECEIVED

By Tracy Crews at 3:37 pm, Jan 03, 2025



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN: 097425 NAME OF AGENCY: St. Louis County Police DATE OF INSPECTION: 01/03/2025

LOCATION OF INSTRUMENT (STREET AND CITY): 14301 South Outer Forty Road, St. Louis, MO 63017 TIME OF INSPECTION: 14:24

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

[X] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

[X] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

[X] PRINTER WORKING PROPERLY

[X] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

[] SIMULATOR SOLUTION

[X] COMPRESSED ETHANOL-GAS MIXTURE

[X] STANDARD SUPPLIER Intoximeters LOT # AG321505 EXP. DATE 08/03/2025

[] SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

[X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [X] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
[] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
[] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .101

TEST 2 = .099

TEST 3 = .099

[X] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS (0-.04) (.05-.09) (.10-.14) (.15-.19) 1 (OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE: [Handwritten Signature]

PRINT NAME: PO Robert Tosie DSN 4553

TYPE II PERMIT NUMBER/EXPIRATION DATE: 240128 / 05/29/2026

TELEPHONE NUMBER: (636) 529-8210

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

OS 10 Serial no: 092425
Version no: 1830

TEST REPORT PRINT

TEST REPORT 09026
of
Date: 09/24/25 Time: 14:01

Dr. Blanks
09/24/25 14:01 090
Calibration Date:
09/24/25 14:01 101

Subject Name

Test #1

Subject I.D.

N/A

Operator Name I.D.

P.O. TOSR 4553

Location

MOBOST TMC

OS 10 Serial no: 092425
Version no: 1830

TEST REPORT 09027
of
Date: 09/24/25 Time: 14:01

Dr. Blanks
09/24/25 14:01 090
Calibration Date:
09/24/25 14:01 101

Subject Name
Test #2

Subject I.D.
N/A

Operator Name I.D.
P.O. TOSR 4553

Location
MOBOST TMC

OS 10 Serial no: 092425
Version no: 1830

TEST REPORT 09028
of
Date: 09/24/25 Time: 14:01

Dr. Blanks
09/24/25 14:01 090
Calibration Date:
09/24/25 14:01 101

Subject Name

Test #3

Subject I.D.
N/A

Operator Name I.D.
P.O. TOSR 4553

Location
MOBOST TMC

OS 10 Serial no: 092425
Version no: 1830

TEST REPORT 09029
of
Date: 09/24/25 Time: 14:01

Dr. Blanks
09/24/25 14:01 090

Subject Name

REF

Subject I.D.
N/A

Operator Name I.D.
P.O. TOSR 4553

Location
MOBOST TMC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 3-Aug-2023

Lot # AG321505 **Model** 108

Exp Date 3-Aug-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
-------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:08.03.2023 17:68

Approved for Release: _____

Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ROBERT TOSIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 572.020 through 577.044, RSMo and 306.111 through 306.119 RSMo.

DATE 5/29/2024

NUMBER 240128

EXPIRES 5/29/2026

Mike Magallon

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dante J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES