



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 7:34 am, Jan 24, 2025

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100289	NAME OF AGENCY MSHP (MSC)	DATE OF INSPECTION 01/21/2025
LOCATION OF INSTRUMENT (STREET AND CITY) 1200 S. Holden St. Warrensburg, MO 64093		TIME OF INSPECTION 8:56 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth LOT # 23390 EXP. DATE 10/17/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP 2116 SIM. NIST EXP DATE 01/13/2026

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .103	TEST 3 .102
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Ryan Schildknecht
TYPE II PERMIT NUMBER/EXPIRATION DATE 230225 10/19/2025	TELEPHONE NUMBER (660) 543-4573

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 01144

Temp Date Time 210L

Air Blank:
01/21/25 08:56 .000
Calibration Check:
21 01/21/25 08:56 .103

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

[Signature] 230225

Location

MSC

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 01145

Temp Date Time 210L

Air Blank:
01/21/25 08:58 .000
Calibration Check:
21 01/21/25 08:58 .103

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

[Signature] 230225

Location

MSC

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 01146

Temp Date Time 210L

Air Blank:
01/21/25 09:00 .000
Calibration Check:
22 01/21/25 09:00 .102

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

[Signature] 230225

Location

MSC

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 01147

Temp Date Time 210L

VOID: RFI
12 01/21/25 09:02

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

[Signature] 23-02-25

Location

MSC



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

RYAN SCHILDKNECHT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT, INTOXILYZER 8000, INTOX EC/IR II, ASIV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/19/2023

NUMBER 230225

EXPIRES 10/19/2025

Laura E. Wag

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 660-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SCHILDKNECHT, RYAN
 Permit No 230225
 Date issued 10/19/2023 Date Expires 10/19/2025



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.